


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220711A/N-118	Page	: 1 of 1	
CALIBRATED ON.	: 12 / 07 / 2022	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 11 / 01 / 2023	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 220711A/N Dt. 11.07.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000084F Dt. 14.07.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001			
ITEM CALIBRATED	: Digital Temp. Indicator of Refrigerator			
	Make : Apteckdeals	Range	: (2 to 8) °C	
	L. Count :	ID No.	: RCL/LAB/TEMP-06	
	Location : Accessioning Dept.	S. No.	:	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022
RESULTS : ET Discipline				
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.10	2	0.10		
4.16	4	0.16		
6.22	6	0.22		
8.28	8	0.28		
(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

