



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220711A/N-102	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 11 / 07 / 2022	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 10 / 01 / 2023	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DATE</b>	: 220711A/N Dt. 11.07.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000068F Dt. 14.07.2022	<b>Cal. At</b>	: Site / Lab

<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001
-----------------------	---

<b>ITEM CALIBRATED</b>	: Micro Pipette
<b>Make</b>	: Fine Care Corporation
<b>L. Count</b>	: Range : (10 to 100) µl
<b>Location</b>	: ID No. : RCL/PIP-02
	: S. No. : QF602118

<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}
-------------------------------	--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100		10	10.10	(-) 0.10
		50	50.20	(-) 0.20
		100	100.32	(-) 0.32

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

**\*U.C.** : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.

Prepared by

Auth. Signatory  
S.K. Sharma  
CEO/ITM

....End of Calibration Certificate....