



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220711A/N-105	Page	: 1 of 1
CALIBRATED ON.	: 11 / 07 / 2022	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 10 / 01 / 2023	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 220711A/N Dt. 11.07.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000071F Dt. 14.07.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001		
ITEM CALIBRATED	: Micro Pipette	Make	: Fine Care Corporation
		Range	: (10 to 100) µl
		L. Count	: ID No. : RCL/PIP-05
		Location	: S. No. : QF602021

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100		10	10.12	(-) 0.12
		50	50.26	(-) 0.26
		100	100.38	(-) 0.38

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmation statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM

....End of Calibration Certificate....