



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 220711A/N-107		Page : 1 of 1		
CALIBRATED ON. : 11 / 07 / 2022		Ref. Std. : CG-11		
CALIBRATION DUE. : 10 / 01 / 2023		C.P. No. : ACS/CP/07		
SRF No. & DATE : 220711A/N Dt. 11.07.2022		Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC342722000000073F Dt. 14.07.2022		Cal. At : Site / Lab		
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001				
ITEM CALIBRATED : Digital Thermo Hygrometer				
Make : HTC		Range : (0 to 50)°C, (10-99)%		
L. Count :		ID No. : RCL/LAB/HGM-01		
Location : Haematology		S. No. :		
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C} R.H. {40% to 60%}				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator.	ACS/RHTIS/01/ 20365872	DVG	CC210121000001169F	15/07/2022
RESULTS : Thermal Discipline				
* U. C. INSTT. READS		STANDARD READING	ERROR	
(°C)	25	25.2	- 0.2	
	28	28.4	- 0.4	
	32	32.6	- 0.6	
(%)	40	39	1	
	50	49	1	
	60	58	2	

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM