




# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220711A/N-108	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 11 / 07 / 2022	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 10 / 01 / 2023	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; DATE</b>	: 220711A/N Dt. 11.07.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272200000074F Dt. 14.07.2022	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001		
<b>ITEM CALIBRATED</b>	: Digital Thermo Hygrometer		
	<b>Make</b> : HTC	<b>Range</b>	: (0 to 50)°C, (10-99)%
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/LAB/HGM-02
	<b>Location</b> : Clinical Biochemistry	<b>S. No.</b>	:
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC21012100001169F	15/07/2022

RESULTS : Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING		ERROR
(°C)	25	25.2	- 0.2
	28	28.2	- 0.2
	32	32.4	- 0.4
(%)	40	38	2
	50	48	2
	60	57	3

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

Auth. Signatory  
S.K. Sharma  
CEO/ITM

....End of Calibration Certificate....