




ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-----------------|
| CERTIFICATE NO. | : 220711A/N-110 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 12 / 07 / 2022 | Ref. Std. | : IS:12508:1988 | |
| CALIBRATION DUE. | : 11 / 01 / 2023 | C.P. No. | : ACS/CP/10 | |
| SRF No. & DATE | : 220711A/N Dt. 11.07.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000076F Dt. 14.07.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. House no. 40-2, Vivek Vihar, Dehradun- 248001 | | | |
| ITEM CALIBRATED | : Centrifuge | | | |
| | Make : Remi | Range | : (0 to 5250) RPM | |
| | L. Count : | ID No. | : RCL/LAB/Cent-02 | |
| | Location : Sample Receiving Room | S. No. | : ZHFN-18244 | |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Tachometer | ACS/DTM/01/ TM5010EK-039 | Godrej | M-210914-16-1 | 14.09.2022 |
| RESULTS | : Mechanical Discipline | | | |
| * U. C. INSTT. READS (RPM) | Standard Measured (RPM) | Error (RPM) | | |
| 1000 | 1003.78 | - 3.78 | | |
| 2000 | 2005.14 | - 5.14 | | |
| 3000 | 3006.50 | - 6.50 | | |
| 4000 | 4006.88 | - 6.88 | | |
| (average of readings) | | | | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

