



CERTIFICATE OF CALIBRATION

Customer Name : Redcliffe Lifetech Pvt. Ltd., Sector-37, Chandigarh

Model : Laura

Serial No. : 0651-1-202106

Calibration Date : 24/01/2022

Calibration Due Date : 23/01/2023

This is to certify that we have calibrated the Automated Urine Chemistry Analyzer, Model: Laura bearing Sr. No. 0651-1-202106 and on inspection of the instrument it is observed that the results are well within the range and instrument is working satisfactorily.

Thanking you,

For **Transasia Bio-Medicals Ltd.**

Rashpal Singh Rana
Sr. Regional Service Manager
M: +91-9316004670
Tel Fax No: +91-172-4643837
E mail: r.rana@transasia.co.in

DEKAPHAN LAURA
Seq.No: 0002
URINORM P
2022.08.04

Mony
Callin
4/8/22
10:44 AM

.....

*BLD	2+
*BIL	3+
*UBG	1+
*KET	2+
*GLU	1+
*PRO	2+
pH	9
*NIT	POS
*LEU	3+
SG	1.000

QC TEST: OK

DEKAPHAN LAURA
Seq.No: 0001
URINORM N
2022.08.04

10:42 AM

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK