



Abbott (Diagnostics Division)

405, KIRTI SHIKHAR BUILDING,
JANAKPURI DISTT. CENTRE,
NEW DELHI - 110 058
TEL : +91 11 25619194 FAX: +91 11 25613392

Dated- 17th Feb 2022

Certificate of Calibration

This is to certify that the below mentioned instrument installed in REDCLIFF LIFETECH Pvt. Ltd Chandigarh has been calibrated by us on 17th June 2022. The next calibration is due on or before 16th June 2023. The values are falling within the specified range.

Instrument Model- **Architect Plus C4000**

Serial No. -**C462211**

Software Version - **9.50**

Bonibir Singh
17/6/22

For Abbott Healthcare Pvt Ltd

Bonibir Singh
Service Engineer
Abbott Diagnostics Division

QC Results List Report

Caro

Operator ID: ADMIN

Unreleased

System serial number: 86547

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
N028/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	ALT	36 U/L		R	05.08.2022 14:58
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	UA	5.9 mg/dL			05.08.2022 12:45
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	GGT	51 U/L			05.08.2022 12:44
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	UHDL	55 mg/dL			05.08.2022 12:44
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	BilD	1.3 mg/dL			05.08.2022 12:44
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	AST	29 U/L			05.08.2022 12:44
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	Urea	39 mg/dL			05.08.2022 12:43
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	Chol	159 mg/dL			05.08.2022 12:43
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	CreaC	1.67 mg/dL			05.08.2022 12:42
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	AlkP	205 U/L			05.08.2022 12:42
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	Phos	4.3 mg/dL			05.08.2022 12:42
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	BiliT	1.6 mg/dL			05.08.2022 12:40
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	GluC	104 mg/dL			05.08.2022 12:39
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	TP	5.4 g/dL			05.08.2022 12:38
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	Trig	100 mg/dL			05.08.2022 12:38
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	CaC	8.7 mg/dL			05.08.2022 12:37
N022/1	1	RANDOX CCLLevel 1	RANDOX CC - Level 1	AlbG	4.1 g/dL			05.08.2022 12:36

[Signature]
05-08-2022

QC Results List Report

C4000

Operator ID: ADMIN

System serial number: 86547

Unreleased

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	UA	9.7 mg/dL			05.08.2022 12:48
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	ALT	110 U/L			05.08.2022 12:48
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	GGT	186 U/L			05.08.2022 12:47
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	UHDL	100 mg/dL			05.08.2022 12:47
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	BiID	2.1 mg/dL			05.08.2022 12:46
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	AST	124 U/L			05.08.2022 12:46
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	Urea	114 mg/dL			05.08.2022 12:46
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	Chol	289 mg/dL			05.08.2022 12:45
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	CreaC	4.65 mg/dL			05.08.2022 12:45
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	AlkP	381 U/L			05.08.2022 12:45
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	Phos	7.4 mg/dL			05.08.2022 12:44
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	BiliT	5.3 mg/dL			05.08.2022 12:43
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	GluC	283 mg/dL			05.08.2022 12:41
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	TP	4.4 g/dL			05.08.2022 12:41
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	Trig	284 mg/dL			05.08.2022 12:40
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	CaC	13.1 mg/dL			05.08.2022 12:40
N022/2	1	RANDOX CCLevel 2	RANDOX CC - Level 2	AlbG	3.1 g/dL			05.08.2022 12:39

 05-08-2022