



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220725A/N-115	Page	: 1 of 1
CALIBRATED ON.	: 26 / 07 / 2022	Ref. Std.	: CG-11
CALIBRATION DUE.	: 25 / 07 / 2023	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 220725A/N Dt. 25.07.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272200000102F Dt. 27.07.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. SCO No171, Sector 37-C, Chandigarh-160036		

ITEM CALIBRATED	: Digital Thermo Hygrometer		
Make	: HTC	Range	: (0 to 50)°C, (10-99)%
L. Count	:	ID No.	: RCL/LAB/HGM-02
Location	: Clinical Pathology	S. No.	:

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Senor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023

RESULTS :			
Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR	
(°C)	25	25.2	- 0.2
	28	28.6	- 0.6
	32	32.8	- 0.8
(%)	40	38	2
	50	48	2
	60	57	3

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth: Signatory
S.K. Sharma
CEO/IM

....End of Calibration Certificate....