


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com 011-2613427



Calibration Certificate

CERTIFICATE NO.	: 220725A/N-106	Page	: 1 of 1
CALIBRATED ON.	: 25 / 07 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 24 / 01 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 220725A/N Dt. 25.07.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272200000093F Dt. 27.07.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. SCO No171, Sector 37-C, Chandigarh-160036		

ITEM CALIBRATED	: Centrifuge		
Make	: Remi	Range	: (0 to 6000) RPM
L. Count	:	ID No.	: CHD/LAB/Cent-01
Location	: Biochemistry	S. No.	: ZHCN-07832

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ TM5010EK-039	Godrej	M-210914-16-1	14.09.2022

RESULTS			
: Mechanical Discipline			
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	-
1000	1003.84	- 3.84	
2000	2004.24	- 4.24	
3000	3006.50	- 6.50	
4000	4006.99	- 6.99	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.88 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.

Prepared by

Authn. Signatory
S.K. Sharma
CEO/TM

....End of Calibration Certificate....