


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 220725A/N-109	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 26 / 07 / 2022	<b>Ref. Std.</b> : ISO:8655:6
<b>CALIBRATION DUE.</b> : 25 / 07 / 2023	<b>C.P. No.</b> : ACS/CP/08
<b>SRF No. &amp; DATE</b> : 220725A/N Dt. 25.07.2022	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342722000000096F Dt. 27.07.2022	<b>Cal. At</b> : Site / Lab

<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. SCO No171, Sector 37-C, Chandigarh-160036
---

<b>ITEM CALIBRATED</b> : Micro Pipette	
Make : Fine Care Corporation	Range : (10 to 100) µl
L. Count :	ID No. : RCL/LAB/PIP-02
Location : Haematology	S. No. : QB578559

<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. { 40% to 60% }
--

<b>STANDARD INSTRUMENTS USED :</b>				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

<b>RESULTS :</b> Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100		10	10.12	(-) 0.12
		50	50.24	(-) 0.24
		100	100.30	(-) 0.30

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.

Prepared by

Adv. Calibration Services  
Auth. Signatory  
S.K. Sharma  
CEO/TM

....End of Calibration Certificate....