


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com • 3427



Calibration Certificate

CERTIFICATE NO. : 220725A/N-112 Page : 1 of 1

CALIBRATED ON. : 26 / 07 / 2022
CALIBRATION DUE. : 25 / 07 / 2023

Ref. Std. : ISO:8655:6
C.P. No. : ACS/CP/08

SRF No. & DATE : 220725A/N Dt. 25.07.2022

Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC34272200000099F Dt. 27.07.2022

Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.
SCO No171, Sector 37-C, Chandigarh-160036

ITEM CALIBRATED : Micro Pipette

Make : Fine Care Corporation Range : (10 to 100) µl
L. Count : ID No. : RCL/LAB/PIP-05
Location : Clinical Biochemistry S. No. : QB578495

ENVIRONMENT CONDITION: Temp. { 25°C + 3°C} R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS : Mechanical Discipline

Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.18	(-) 0.18
100		50	50.28	(-) 0.28
		100	100.42	(-) 0.42

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.6 \mu\text{l}$ at coverage factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

