


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com **CC-3427**



Calibration Certificate

CERTIFICATE NO.	: 220730A/NN-105	Page	: 1 of 1	
CALIBRATED ON.	: 30 / 07 / 2022	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 29 / 07 / 2023	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 220730A/NN Dt. 30.07.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000146F Dt. 31.07.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. SCO No. 171, Sector 37-C, Chandigarh-160036			
ITEM CALIBRATED	: Digital Temp. Indicator of Refrigerator			
	Make : CE	Range	: (2 to 8) °C	
	L. Count :	ID No.	: CHD/Lab/D.Temp-05	
	Location : Sample Storage Fridge	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022
RESULTS	: ET Discipline			
	* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
	2.05	2	0.05	
	4.10	4	0.10	
	6.14	6	0.14	
	8.18	8	0.18	
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

