



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 4751

Distribution No.: 155-L

Month/Year: April/2022

Instrument ID: 06496

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 04-07-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	12.4	11.5	23.9	24.12	0.1270	-0.07	0.9	0.2	0.0130	3.07
RBC x10 <sup>6</sup> /µl	1	5.92	5.78	11.7	12.05	0.0170	-0.76	0.14	0.06	0.0040	1.35
Hb g/dl	1	15.1	14.7	29.8	28.5	0.0290	1.59	0.4	0.1	0.0080	4.05
HCT%	1	45.9	45.2	91.1	95.8	0.2590	-0.62	0.7	0.5	0.0360	0.34
MCV-fl	1	78.2	77.5	155.7	159.6	0.3180	-0.43	0.7	0.3	0.0210	1.08
MCH-Pg	1	25.5	25.4	50.9	47.1	0.0570	2.44	0.1	0.2	0.0120	-0.67
MCHC-g/dl	1	32.9	32.5	65.4	58.9	0.1470	1.57	0.4	0.25	0.0170	0.51
Plt. x10 <sup>3</sup> /µl	1	206	202	408	399	2.71	0.11	4	9	0.55	-0.52
Retic %	2	3.5	3	6.5	17.05	0.44	-0.77	0.5	0.7	0.05	-0.16

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=3 , Poly=59 L=3, E=5, Mono/Promono=1 , B1=1 P.M.=1, Mye=22, Meta=7, Other=0
RBC Morphology	3	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
Diagnosis	3	CHRONIC MYELOID LEUKEMIA CHRONIC PHASE

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 155--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	347	334	87.43	89.22	5.69	4.19	6.88	6.59
RBC x10 <sup>6</sup> /µl	1	347	347	83.29	86.46	7.49	2.02	9.22	11.52
Hb g/dl	1	347	347	87.32	84.73	4.9	6.05	7.78	9.22
HCT%	1	347	333	90.39	88.89	6.61	5.11	3	6
MCV-fl	1	347	333	90.09	86.19	5.41	3.9	4.5	9.91
MCH-Pg	1	347	333	87.99	93.69	7.51	3	4.5	3.31
MCHC-g/dl	1	347	333	91.59	86.79	4.8	4.5	3.61	8.71
Plt. x10 <sup>3</sup> /µl	1	347	333	96.7	86.19	1.8	8.71	1.5	5.1
ReticCount%	2	347	222	91.44	93.24	5.41	2.7	3.15	4.06
PS Assessment	3	347	230	Satisfactory :96.26%, Borderline Sat. :2.88%, Unsatisfactory :0.86%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----