



PC - 1034

Indian Association of Medical Microbiologists
Virology External Quality Assessment Scheme
CMCVIROEQAS, Proficiency Testing Unit, Department of Clinical Virology
Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455
16th SEROLOGY CMCVIROEQAS EVALUATION FINAL REPORT

PANEL: Dengue Serology

CMCVIROEQAS ID.:

Opening Date: 21/03/2022

Result Receiving Date: 11/04/2022

Distribution No:

Specimen #	Intended Result			Your Result		
	IgM	IgG	NS1	Dengue IgM	Dengue IgG	Dengue NS1
S02122013	Negative	Negative	Negative	NEGATIVE	NEGATIVE	NEGATIVE
S02122014	Negative	Negative	Positive	NEGATIVE	NEGATIVE	POSITIVE
S02122015	Positive	Positive	Negative	POSITIVE	POSITIVE	NEGATIVE
S02122016	Negative	Negative	Negative	NEGATIVE	NEGATIVE	NEGATIVE
Your Score				8/8 (100%)	6/6 (100%)	8/8 (100%)
Performance grading				E	E	E

DNP- Did not participate

Scoring System

Qualitative Results	Score
Concordant Result	2
Intermediate/Indeterminate	1
Discordant Result	0



Indian Association of Medical Microbiologists
Virology External Quality Assessment Scheme
CMCVIROEQAS, Proficiency Testing Unit, Department of Clinical Virology
Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

PC - 1034

16th SEROLOGY CMCVIROEQAS EVALUATION FINAL REPORT

PANEL: BBVS

CMCVIROEQAS ID. V1014

Opening Date: 21-03-2022

Result Receiving Date: 10/04/2022

Distribution No: S122

Specimen #	Intended Result			Your Result		
	HIV	HBsAg	HCV	HIV	HBsAg	HCV
S01122001	Negative	Positive	Negative	NEGATIVE	POSITIVE	NEGATIVE
S01122002	Negative	Negative	Negative	NEGATIVE	NEGATIVE	NEGATIVE
S01122003	Negative	Negative	Negative	NEGATIVE	NEGATIVE	NEGATIVE
S01122004	Negative	Negative	Positive	NEGATIVE	NEGATIVE	POSITIVE
Your Score				8/8 (100%)	8/8 (100%)	8/8 (100%)
Performance grading				E	E	E

Scoring System

Qualitative Results	Score
Concordant Result	2
Intermediate/Indeterminate	1
Discordant Result	0



Indian Association of Medical Microbiologists
Virology External Quality Assessment Scheme
CMCVIROEQAS, Proficiency Testing Unit, Department of Clinical Virology
Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

PC - 1034

11th MOLECULAR CMCVIROEQAS EVALUATION FINAL REPORT

PANEL: HCV RNA Quantitative

CMCVIROEQAS LAB ID: **V1014**

Opening Date: 21/03/2022

Result Receiving Date: 06/04/2022

CMCVIROEQAS Distribution No: **M122**

Specimen #	Intended Result	Your Result
M03122069	2.9 log IU/ml	NOT DETECTED
M03122070	4.6 log IU/ml	0.8 log IU/ml

Difference in log IU/mL between specimen ID M03122069 and M03122070 is 1.7 Log IU/mL

Your laboratories difference in log IU/mL could not be assessed

Your score is 0/4

Scoring System

Quantitative Results	Score
Within 0.5 Log IU/mL difference between samples	4
Within 0.5 to 1 Log IU/mL difference between samples	2
Greater than 1 Log IU/mL difference between samples	0