

CALIBRATION CERTIFICATE

To,

This is to Certify that Biochemistry Analyzer Meril **CliniQuant Micro** Instrument Sr.No. **CM0921150** Installed at **Renuka Diagnostics Pathology - Bilaspur** Add: 353, Near Maharana Pratap Chowk, Ring Road No. 2, Gauvrav Path, Bilaspur, Chattisgarh. 495001 _ has been calibrated by our engineer on **02-Aug -2022** with appropriate and sophisticated tools. Bellow parameters are been calibrated:

TEST ITEMS AND RESULT

Appearance	OK
AD Value Range (120<AD<350)	OK
Temperature Checking (at 37'C)	OK
Printer Test	OK
Aspiration Volume Range	OK
Stability of Pot. Dichromate	OK
Absorbency Linearity at 340	OK
PC Communication Checking	OK
Sample Test 1	OK
Sample Test 2	OK
Sample Test 3	OK

This Calibration Certificate is Valid Up to **01-Aug-2023** .

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

Engineer Name - Mr. Mayank Vaishnav.

Designation -Service Engineer.

Thanks and Regards,



For Meril Diagnostics Pvt. Ltd.

Meril Diagnostics Pvt. Ltd.

Corporate Office: Survey No. 135/ 139, Bilakhia House, Muktanand Marg, Chala, **Vapi** – 396191, **Gujarat**

Mumbai Office: 602, Midas, Sahar Plaza, JB Nagar, Andheri East, **Mumbai** – 400059, **Maharashtra**

Call Centre: 18002663745 Direct Line: +91-22-30732544 E mail: diagnostics.csd@merillife.in

T +91 260 3052100
F +91 260 3052125

www.merillife.com

Meril Diagnostics Pvt. Ltd.
Survey No.135/2/B, Bilakhia House
Muktanand Marg, Chala
Vapi - 396191. Gujarat, India.

QUALITY CERTIFICATE

To,

This is to Certify that Biochemistry Analyzer Meril **CliniQuant Micro** Instrument Sr.No. **CM0921150** Installed at **Renuka Diagnostics Pathology - Bilaspur** Add: 353, Near Maharana Pratap Chowk, Ring Road No. 2, Gauvrav Path, Bilaspur, Chattisgarh. 495001 has been Quality Control by our engineer on **02-Aug -2022** with appropriate and sophisticated tools. Bellow parameters are been qualified:

TEST ITEMS AND RESULT

Sr. No.	Product Code	Material Description	Pack Size (ml)	Result
1	ALBFSR-01	ALBUMIN	4 x 50	3.55
2	BILFSR-01	BILIRUBIN (TOTAL & DIRECT)	4 x 50	T-1.35, D-0.90
12	HDCFSR-01	HDL - CHOLESTEROL	4 x 25	40.0
13	CREFSR-03	CREATININE	2 x 25 / 2 x 25	1.4
20	ALPFSR-01	ALKALINE PHOSPHATASE	4 x 10	72.10
22	AMYFSR-01	AMYLASE	4 x 10	90.38
23	GLUFSR-02	GLUCOSE	5 x 100	93.50
27	CHOFSR-01	CHOLESTEROL	4 x 25	138.8
30	GPTFSR-01	SGPT	4 x 20 / 4 x 5	44.24
33	GOTFSR-01	SGOT	4 x 20 / 4 x 5	41.42
34	UREFSR-01	UREA	4 x 20 / 4 x 5	17.70
36	URCFSR-01	URIC ACID	4 x 10	5.30
43	TPRFSR-01	TOTAL PROTEIN	4 x 50	5.80
45	TRGFSR-02	TRIGLYCERIDE	4 x 50	110.0
47	LIPFSR-01	LIPASE	1 x 16/1 x 10	45.7
48	HDLFSR-01	HDL CHOLESTEROL Direct	2 x 24 / 2 x 8	40
50	LDLFSR-01	LDL CHOLESTEROL - DIRECT	1 x 24 / 1 x 8	80

This Quality Certificate is Valid Up to **01-Aug-2023** .

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

Engineer Name – Mr. Mayank Vaishnav.

Designation –Service Engineer.

Thanks and Regards,



For Meril Diagnostics Pvt. Ltd.

Meril Diagnostics Pvt. Ltd.

Corporate Office: Survey No. 135/ 139, Bilakhia House, Muktanand Marg, Chala, **Vapi – 396191, Gujarat**

Mumbai Office: 602, Midas, Sahar Plaza, JB Nagar, Andheri East, **Mumbai – 400059, Maharashtra**

Call Centre: 18002663745 Direct Line: +91-22-30732544 E mail: diagnostics.csd@merillife.in

T +91 260 3052100
F +91 260 3052125

www.merillife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135/2/B, Bilakhia House
Muktanand Marg, Chala
Vapi - 396191, Gujarat, India.