



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3169

Distribution No.: 155-H

Month/Year: March/2022

Instrument ID: SN 1740707210744

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 11-05-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	3.17	3.12	6.29	8.1	0.0550	-1.99	0.05	0.1	0.0090	-0.42
RBC x10 <sup>6</sup> /µl	1	3.33	3.31	6.64	6.8	0.0100	-0.92	0.02	0.04	0.0030	-0.60
Hb g/dl	1	11	11	22	24	0.0400	-2.70	0	0.1	0.0110	-0.67
HCT%	1	33.7	33.5	67.2	70.1	0.1840	-0.81	0.2	0.4	0.0340	-0.54
MCV-fl	1	101.3	101.2	202.5	207.3	0.4150	-0.65	0.1	0.4	0.0460	-0.54
MCH-Pg	1	33.2	33	66.2	70.1	0.1370	-1.62	0.2	0.3	0.0260	-0.45
MCHC-g/dl	1	32.8	32.6	65.4	68	0.2000	-0.66	0.2	0.3	0.0340	-0.27
Plt. x10 <sup>3</sup> /µl	1	141	134	275	294	1.69	-0.57	7	5	0.49	0.39
Retic %	2	5.5	5	10.5	12.5	0.43	-0.23	0.5	0.5	0.05	0.00

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=0 , Poly=53 L=5, E=1, Mono/Promono=2 , B1=7 P.M.=2, Mye=24, Meta=5, Other=
RBC Morphology	3	Poly: 38 - 52, Myelo: 14 - 25, Meta: 7 - 16, Blast: 2-8, Lympho: 2-6 , Promyelo: 1-5 nRBC/Eos/Baso/Mono: 0 - 5
Diagnosis	3	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
		PBS Findings S/O Chronic Myeloproliferative Disorder CMPD (?CML), Advice- Bone Marrow Examination ,Philadelphia Chromosome
		Chronic Myeloid Leukemia (Chronic Phase)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 155--H	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	151	151	84.11	91.39	1.32	1.32	14.57	7.29
<b>RBC x10<sup>6</sup>/µl</b>	1	151	151	86.75	88.74	9.27	7.28	3.98	3.98
<b>Hb g/dl</b>	1	151	151	90.07	90.07	7.28	2.65	2.65	7.28
<b>HCT%</b>	1	151	151	93.38	90.73	5.96	5.96	0.66	3.31
<b>MCV-fl</b>	1	151	151	91.39	94.04	6.62	3.97	1.99	1.99
<b>MCH-Pg</b>	1	151	151	91.39	88.74	4.64	5.96	3.97	5.3
<b>MCHC-g/dl</b>	1	151	151	94.04	94.7	5.3	2.65	0.66	2.65
<b>Plt. x10<sup>3</sup>/µl</b>	1	151	151	96.69	90.73	2.65	5.3	0.66	3.97
<b>ReticCount%</b>	2	151	119	90.76	92.44	5.88	11.76	3.36	-4.2
<b>PS Assessment</b>	3	151	121	Satisfactory :86.1%, Borderline Sat. :10.59%, Unsatisfactory :3.31%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----