

**Result Entry Report**

Lab : 640 / SP HITECH PATHOLOGY LABORATORY

Cycle No : 220104

Sample No : 01

**Analyte Scheme / Group :** Clinical Path - Urine Routine Chemistry

**Analyte :** pH

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	5	5.5	-

**Analyte :** Specific Gravity

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1.025	1.025	-

**Analyte :** Glucose (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	3+	3+	-

**Analyte :** Protein (Urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1+	2+	-

**Analyte :** Ketone (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Negative	Negative	-

**Analyte :** Blood (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1+	2+	-

**Analyte :** Bilirubin (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Negative	Negative	µmol/L

**Analyte :** Urobilinogen (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Normal	Normal	-

**Result Entry Report**

Lab : 640 / SP HITECH PATHOLOGY LABORATORY

Cycle No : 220104


Sample No : 01

**Analyte : Nitrite (urine)**

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Absent	Absent	-

**Analyte : Urine Pregnancy Test**

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Immunochemistry	-	06/05/2022	Negative	Negative	-



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SP HITECH PATHOLOGY LABORATORY (640)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

**Instrument : -**


Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Negative	Neg

**Instrument : Uriplus 200**

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5.5	5,5.5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,Tr
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	2+	+1,+2,Tr
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Absent	Abs

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	0
Evaluated Parameters	10
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report