



# GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>SP HITECH PATHOLOGY LABORATORY</b> Trimurti Plaza,Near Pavana Sahakari Bank, Somatane Phata Talegaon Dabhade Somatane Phata, Talegaon Dabhade, PCMC, Pune, Maharashtra 410506	Page No. :- 1 of 1 Certificate No. :- GTS/220527/02- 010 Date of Received :- 25.05.2022 Date of Calibration :- 25.05.2022 Next Calibration Due On :- 24.05.2023 Issue Date :- 27.05.2022 Calibration method No. :- MECH-WI-15
Ambient Temp. (°C) :- 21 TO 40 Relative Humidity (%RH) :- 30 TO 75 Location of calibration :- On Site Condition of Item :- Ok	

### 2. Description of Item

Name :- Microscope	Range :- 4x, 10x, 40x 100x
Id No :- SPHL/EQ/MSP/14	Least Count :- - X
Make :- LABOMED	Location Director cabin
Type -	Sr No --

### 3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Linear Glass Scale	FCS/2021-22/17-004	Fine Calibration Services	GTS/ LGS-01	25.05.2022

### \*Mechanical Calibration

#### 4.Calibration Results :-

Calibration Points	Standard Reading	UUC Reading	Error in	Expanded Uncertainty in ± %
X	X	X	X	
4	4	4	0	0.25
10	10	10	0	0.25
40	40	40	0	0.25
100	100	100	0	0.25

#### Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

Rohit

Calibration Engineer

Rohit.L



Approved By

Swarnil

Technical Manager

Swapnil Bhagwat

RF-55/01

End of Certificate

Email : globaltechnical007@gmail.com