



Lab Code No. 2046

BASIC SEROLOGY
FINAL RESULT ASSESSMENT
Amended Report

CYCLE NO. : 11 ROUND: 1 TOTAL PARTICIPANT: 294 DATE: 21/02/2022

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
S1- HIV	256	Non-Reactive	Reactive : 0 Non-Reactive : 256	100%	Within Consensus
S2- HBsAg	258	Reactive	Reactive : 248 Non-Reactive : 10	96%	Within Consensus

Note-Your result is not included in the consensus calculation due to late submission of result.

Revised result received. Results within consensus

Chief Coordinator
Director

Programme

[Signature]
Dr. Sanjay Mehrotra

Reviewed
[Signature]
06/08/2022

[Signature]
Dr. Bandana Mehrotra

Checked By:

End of Report

Prepared by: NS

sunita more

Subject: FW: Regarding reissue of result

From: RML [mailto:rmresearchfoundation@gmail.com]
Sent: Saturday, August 06, 2022 11:30 AM
To: Dr. Pushkar Matkari
Subject: Re: Regarding reissue of result

Dear Participant,

Thanks for your email.

You have participated in our Serology EQAS Program Cycle 11 Round 1. The result submission date was 01.02.2022, but you have submitted

your result on 07.03.2022 due to which your result was not included in the consensus calculation.

Looking at your result for **HIV Sample**, your result reporting was **Non-Reactive** as it is in concordance with 100% consensus

and for your sample 2 for **HBsAg**, your result reporting was **Reactive** and it is concordance with the 96% consensus.

You are requested to submit the result on time so that your result are include in consensus calculation.

Thanks & Regards,

RML-QAP

Contact. No:- 0522-4034100

Ph.no-7518077222

NOTE; We will be available Monday to Saturday- 9.AM TO 7.PM