


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 220801A/N-109	Page	: 1 of 1
CALIBRATED ON.	: 01 / 08 / 2022	Ref. Std.	: DKD R-5-7
CALIBRATION DUE.	: 31 / 07 / 2023	C.P. No.	: ACS/CP/11
SRF No. & DATE	: 220801A/N Dt. 01.08.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000164F Dt. 02.08.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 12/475, Ultra Tower, Mrobertgaj Near Shani Dev Mandir Chunniganj, Kanpur, Kanpur Nagar U.P.		
ITEM CALIBRATED	: Digital Thermo Hygrometer		
	Make : HTC	Range	: (0 to 50)°C, (10-99)%
	L. Count :	ID No.	: RCL/LAB/HGM-03
	Location : Sample Collection Room	S. No.	: N.A

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023

RESULTS : ET Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR	
(°C)	25	25.3	- 0.3
	28	28.5	- 0.5
	32	32.6	- 0.6
(%)	40	38	2
	50	47	3
	60	57	3

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4°C, 1.5% at coverage Factor k=2.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

