



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|-------------------------------------|------------------------|--------------|
| CERTIFICATE NO. | : 220801A/N-102 | Page | : 1 of 1 |
| CALIBRATED ON. | : 01 / 08 / 2022 | Ref. Std. | : ISO:8655:6 |
| CALIBRATION DUE. | : 31 / 01 / 2023 | C.P. No. | : ACS/CP/08 |
| SRF No. & DATE | : 220801A/N Dt. 01.08.2022 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342722000000157F Dt. 02.08.2022 | Cal. At | : Site / Lab |

| | |
|-----------------------|--|
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. 12/475, Ultra Tower, Mrobertgaj Near Shani Dev Mandir Chunniganj, Kanpur, Kanpur Nagar U.P. |
|-----------------------|--|

| | |
|------------------------|----------------------------|
| ITEM CALIBRATED | : Micro Pipette |
| Make | : Fine Care Corporation |
| L. Count | : Range : (100 to 1000) µl |
| Location | : ID No. : RCL/LAB/PIP-02 |
| | : S. No. : QJ615584 |

| | |
|-------------------------------|--|
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. {40% to 60%} |
|-------------------------------|--|

| STANDARD INSTRUMENTS USED : | | | | |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/ D449927064 | BlueBox | BBT/223/JAN/22 | 06/01/2023 |

| RESULTS : Mechanical Discipline | | | | |
|---------------------------------|-----------------|--------------------|----------------------|------------|
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 100 | 100.34 | (-) 0.34 |
| 1000 | | 500 | 501.60 | (-) 1.60 |
| | | 1000 | 1003.24 | (-) 3.24 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida

....End of Calibration Certificate....