


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

|                                 |                                     |                        |              |
|---------------------------------|-------------------------------------|------------------------|--------------|
| <b>CERTIFICATE NO.</b>          | : 220801A/N-106                     | <b>Page</b>            | : 1 of 1     |
| <b>CALIBRATED ON.</b>           | : 04 / 08 / 2022                    | <b>Ref. Std.</b>       | : ISO:8655:6 |
| <b>CALIBRATION DUE.</b>         | : 31 / 01 / 2023                    | <b>C.P. No.</b>        | : ACS/CP/08  |
| <b>SRF No. &amp; DATE</b>       | : 220801A/N Dt. 01.08.2022          | <b>Phy. Inspection</b> | : O.K.       |
| <b>ULR No. &amp; ISSUE DATE</b> | : CC342722000000161F Dt. 02.08.2022 | <b>Cal. At</b>         | : Site / Lab |

|                       |  |
|-----------------------|--|
| <b>CALIBRATED FOR</b> | : M/s Redcliffe Lifetech Pvt. Ltd.<br>12/475, Ultra Tower, Microbertgaj Near Shani Dev Mandir<br>Chunniganj, Kanpur, Kanpur Nagar U.P. |
|-----------------------|--|

|                        |                         |        |                    |
|------------------------|-------------------------|--------|--------------------|
| <b>ITEM CALIBRATED</b> | : Micro Pipette         |        |                    |
| Make                   | : Dragon Lab            | Range  | : (100 to 1000) µl |
| L. Count               | :                       | ID No. | : RCL/LAB/PIP-06   |
| Location               | : Clinical Biochemistry | S. No. | : YE193AL0316131   |

|                               |  |
|-------------------------------|--|
| <b>ENVIRONMENT CONDITION:</b> | Temp. { 25°C ± 3°C } R.H. { 40% to 60% } |
|-------------------------------|--|

| STANDARD INSTRUMENTS USED : |                       |              |                 |            |
|-----------------------------|-----------------------|--------------|-----------------|------------|
| INSTT. NAME                 | ID No. / S. No.       | TRACEABILITY | CERTIFICATE No. | DUE DATE   |
| Weighing Balance            | ACS/DB/02/ D449927064 | BlueBox      | BBT/223/JAN/22  | 06/01/2023 |

| RESULTS : Mechanical Discipline |                 |                    |                      |            |
|---------------------------------|-----------------|--------------------|----------------------|------------|
| Nominal Capacity (µl)           | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| 1000                            |                 | 100                | 100.40               | (-) 0.40   |
|                                 |                 | 500                | 501.52               | (-) 1.52   |
|                                 |                 | 1000               | 1003.26              | (-) 3.26   |

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

Auth. Signatory  
  
S.K. Sharma  
CEO/TM  
C-23, Sec-10, Noida