

Daily QCResult

QC Name: QC NORM

QC Lot No.: 30080292

Item Name: ALB

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	4.05	0.30	3.89	1	2022-08-07	
						Total:1

Item Name: ALP-IFCC

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	104	11.5	89.96	1	2022-08-07	< -1SD
						Total:1

Item Name: ALT

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	48.6	4.5	44.17	1	2022-08-07	
						Total:1

Item Name: AST

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	38.0	3.5	37.76	1	2022-08-07	
						Total:1

Item Name: CAL

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	8.50	0.46	8.15	1	2022-08-07	
						Total:1

Item Name: CHOL

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	174.0	11.5	169.6	1	2022-08-07	
						Total:1

Item Name: DBIL

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	0.14	0.055	-0.03	1	2022-08-07	< -3SD
						Total:1

Item Name: E-CRE

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	0.90	0.125	0.88	1	2022-08-07	
						Total:1

Item Name: GLU

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	81.00	6.25	77.82	1	2022-08-07	
						Total:1

Item Name: HDL-C

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	54.60	4.5	54.6	1	2022-08-07	
						Total:1

Item Name: TBIL

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	1.00	0.15	0.77	1	2022-08-07	< -1SD
						Total:1

Item Name: TGL

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	104.00	9.5	102.8	1	2022-08-07	
						Total:1

Item Name: TP

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	6.40	0.475	5.9	1	2022-08-07	< -1SD
						Total:1

Item Name: UA

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	5.20	0.45	4.9	1	2022-08-07	
						Total:1

Item Name: UREA

Sample Type	Target Value	Standard Deviation	Result	Times	Date	Status
Serum	26.00	3.9	27.86	1	2022-08-07	
						Total:1

**CERTIFICATE OF CALIBRATION**

ADL/DO/MS/C/0145/22-23

Instrument: Mispa CXL PRO Fully Automated Biochemistry Analyser**Customer Name:** Dr. Vaibhav's Pathocity Path Lab & Imaging, Paschim Vihar, New Delhi**Date of Installation:** 27/Aug/2021**S/N:** 210T180CS0218K**Alignment Check**

Motion Check	Checked	Remark
Sample Probe	<input checked="" type="checkbox"/>	OK
Reagent Probe	<input checked="" type="checkbox"/>	OK
Mixing Bar	<input checked="" type="checkbox"/>	OK
Reaction Disk	<input checked="" type="checkbox"/>	OK
Reagent Disk	<input checked="" type="checkbox"/>	OK
Sample Disk	<input checked="" type="checkbox"/>	OK
Cuvette Check	<input checked="" type="checkbox"/>	OK
Lamp	<input checked="" type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	OK
Temp. unit	<input checked="" type="checkbox"/>	OK
Mixing unit	<input checked="" type="checkbox"/>	OK
Reaction unit	<input checked="" type="checkbox"/>	OK
Reagent unit	<input checked="" type="checkbox"/>	OK
Sample unit	<input checked="" type="checkbox"/>	OK

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

OPTICAL SECTION

Filter	Light Intensity	
340	14072	
380	14200	
405	13720	
450	12712	
480	12100	
505	12321	
546	12057	
570	11823	
600	11321	
660	11543	
700	11357	
800	11052	

Power Supply : 12V 24V

The equipment is calibrated for the different parameters by using Multi Calibrator

Lot.No: 32010036 Exp: 2023-10

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the Agappe Diagnostics Ltd.

Date of Calibration : 26/Feb/2022

Next Due : 25/Aug/2022

For Agappe Diagnostics Ltd.,




Authorized Signatory

Name: Jitender Kumar

Designation: Asst. Manager – Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nelliad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in