



Calibration Certificate

This is to Certify that Biochemistry Analyzer Meril **AutoQuant-200i** Instrument Sr. No. **AQ-050974** Installed at **Delhi Clinical Laboratory, Haryana** has been calibrated by our engineer on **07-Jun-2022** with appropriate and sophisticated tools. Below parameter has been calibrated:

TEST ITEMS AND RESULT

Appearance	No Imbalance Probe Movement
Leakage in Water tubing or CRU	No water leakage
Ambient temp (15- 32) Degree	24.0 degree C
Lamp Check	Ok
Sample Dispersing Volume	In range (0 to 30 ul)
Reagent cooling Temperature	2.8 ° C (0 to 15 Degree)
Optical Path length	0.65
Waster & Water Sensor	Working Properly
Threshold value	±34000 (≥25,000)
Valve & Motor	Working Properly
Reagent Dispensing Volume	In range (0 to 150 ul)
Power Supply and Earthing	230 Volts and Grounding less then 5 V
CV-Kinetic/End point/two Point	CV- 1.77% /2.42% /2.80%
SD-Kinetic/End point/two Point	SD- 1.15/0.39/0.69
Mean-Kinetic/End point/two Point	Mean- 97.55/14.34/24.79
Hardware Check-Functional Test	OK5415152524141112
Light Quality Check	OK
Cuvette Absorbance	OK
Cuvette Blank Value	OK >33000
Incubation temp.	37.0 ° C
Room temp.	25.0 degree C

This Calibration Certificate is Valid Up to **07-Jun-2023**.

Conclusion: All parameter are within range. The Machine is working properly.
(As per Meril terms.)

Engineer Name: **Mr. Arun Yadav**

Designation: **Service Engineer**

Engineer Sign:




Delhi Clinical Laboratory - Haryana

Installation Qualification

Operational Qualification

Performance Qualification



AutoQuant 200i

Fully Automated Biochemistry Analyzer



T +91 260 3052100
F +91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 126, 2nd, 3rd & 4th floors
Muktanand Marg, Chola
Vapi - 396191, Gujarat, India

Cardiovascular

Orthopedic

Surgical

Page 1 of 37



AutoQuant 200i
Fully Automated Biochemistry Analyzer

INSTALLATION QUALIFICATION

FOR

Delhi Clinical Laboratory - Haryana

INSTALLATION QUALIFICATION#01

MARKETED BY:
MERIL DIAGNOSTICS PVT. LIMITED
MUKTANAND MARG, CHALA
VAPI - 396191
INDIA

T +91 260 3052100
F +91 260 3052125
www.merilife.com



Meril Diagnostics Pvt. Ltd.
Muktanand Marg, Chala
Vapi - 396191

TABLE OF CONTENTS

Sr. No. CONTENTS

- I. Approval of the IQ procedure
- II. Instructions
- III. Scope
- IV. Ancillary Information
- V. Installation Qualification
- VI. Installation procedure
- VII. Comments
- VIII. System Certification

Appendices

- i. IQ/OQ/PQ dot.
- ii. Sample page of the log book
- iii. Company Certificate



T +91 260 3052100
F +91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 109, 1st Phase, H. H. S.
Muktanand Marg, Chula
Vapi, 396191, Gujarat, India

Cardiovascular

Orthopedic

Surgical

Page 3 of 37

I. APPROVAL OF THE IQ PROCEDURE:

Meril Diagnostics Pvt. Ltd. is responsible for the installation of the **Fully Automatic Biochemistry Analyzer, Merilyzer - AutoQuant 200i, Sr. No. AQ-050974** at **Delhi Clinical Laboratory - Haryana** as per the attached protocol.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Title: INSTALLATION QUALIFICATION

Company: MERIL DIAGNOSTICS PVT. LIMITED

Signature: 

Date: 07-Jun-2022

Validation Team from CPU:

Name: Mr. Rahul Saini


Title: DRSM – Technical & Scientific Support

Department: Applications & Scientific Support

Signature: 

Date: 07-Jun-2022

Customer Authorizations:

Name : 

Signature: 

Title

Date: 07/20/2022

T : +91 260 3052100

F : +91 260 3052125

www.merilife.com



Meril Diagnostics Pvt. Ltd.
Survey No. 125 - B, Sector 14, Gurgaon
Mustafa Road, Meril, Gurgaon
Vap: 30650 - Gurgaon, Haryana

II. INSTRUCTIONS

This document is to be completed at the time is shifted to its current location (new) and set up for operation.

1. An authorized Meril Diagnostics representative will check the system and enter the specific data as outlined in the appropriate Installation Qualification.
2. Employees of (customer) Delhi Clinical Laboratory - Haryanawill verify each result and sign in the last page. The members of the validation will carry this out.
3. All deviation from the normal specification to include any problems with installation will be noted under COMMENTS. All resolution to such problems will also be noted in the COMMENTS section. Additional space is provided at the end of this protocol for the same.
4. This document propriety information and is in no way to be copied, photographed or duplicated in any way without expressed authorization from Meril Diagnostics Pvt. Limited.

Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM – Technical & Scientific Support

Signature: 

Date: 07-Jun-2022

Tel: +91 (0) 30 2 30 2 100

Fax: +91 (0) 30 2 30 2 120

www.merilife.com



Meril Diagnostics Pvt. Limited
Survey No. 10/1, Sector 10, Gurgaon
Mustafabad, Gurgaon, Haryana
Vapi - 396 130, Gujarat, India

III. SCOPE

This installation Qualification protocol will be performed on the **Fully Automatic Biochemistry Analyzer, Merilyzer - AutoQuant 200i, Sr. No. AQ-050974** located at Delhi Clinical Laboratory - Haryana. This protocol will define the documentation that will be used to evaluate the instrument installation in accordance with the manufacturer's specifications and intended use.

Successful completion of this protocol will verify that this instrument has been installed in accordance with the intended usage.


Installation checks will also be performed to verify that the instrument has been installed with proper connections and utilities.

Trained, knowledgeable personnel will perform qualification studies.

Any exceptional condition encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Name: Mr. Rahul Saini

Designation: DRSM – Technical & Scientific Support

Signature: 

Date: 07-Jun-2022



T : +91 360 3052100
F : +91 360 3052125
www.meril.co.com

Meril Diagnostics Pvt. Ltd.
Survey No. 10/1, Block 10
Mushayana, Meerut, Uttar Pradesh
India - 201001, Tel: +91 360 3052100



Design solutions


IV. ANCILLARY INFORMATION.

a. Certificate of Purchase Order Compliance

I certify to the best of my knowledge, the instrument installed on 07-Jun-2022, has been purchased under purchase order No. _____ dated _____ OR contracted for period of *** years is in compliance with the specifications of the purchase order or contract.

Verified By:

Designation: DRSM – Technical & Scientific Support

Signature: 

Date: 07-Jun-2022



Tel: +91 260 3052100
Fax: +91 260 3052125
www.meril.com

Meril Design Solutions Pvt. Ltd.
Survey No. 10/1, Block 1, Phase 1, IIT
Mumbai, Maharashtra
Vap. 400 09, District, MH09



Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature:

Date: 07-Jun-2022

V. INSTALLATION QUALIFICATION

a. Equipment Description

The Automated Biochemistry Analyzer is a walk away system with simultaneous processing clinical chemistry samples. The data files created can be printed in report form or stored as data files for a variety of clinical chemistry-based applications.

	Verified By	Date
Instrument Identification		
Equipment Name : Automated Biochemistry Analyzer	Mr. Arun Yadav	07-Jun-2022
Model : AutoQuant 200i		
Manufacturer : Meril Diagnostics Pvt. Ltd. India		
Marketed By : MERIL DIAGNOSTICS PVT. LIMITED		
Equipment # : Biochemistry Analyzer		
Serial Number : AQ-050974		
Size : 548 X 422 X 383 (mm)		
Power : AC 230 V		



T : +91 260 3052100
F : +91 260 3052125
www.merildiag.com

Meril Diagnostics Pvt. Ltd.
Survey No 135/1B, 1st Floor,
Mukund Marg, Chhala
Vijay Nagar, Gurgaon, Haryana

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature: 

Date: 07-Jun-2022

b. Accessories/Consumables

The following accessories were supplied with the instrument. Check 'verified by' in case they are found to be in order.

Description	Quantity	Verified by	Date
User's manual	1	Mr. Arun Yadav	07-Jun- 2022
Application Software	1		
CPU+Monitor+Keyboard+Mouse+Speakers+Printer	1		
Reagent Vials (20 ml)	50		
Distilled Water Can	1		
Waste Can	1		
Sample Cups	200		
RS 232 Cable	1		
RS 232 to USB Cable	1		
Sample Racks	1		
Reagent Rack	1		
Probe Cleaning Wire	1		

T : +91 260 3052100

F : +91 260 3052125

www.merilife.com



Meril Diagnostics Pvt. Ltd.
Survey No. 125, B-Block, Sector
Mustafabad, Mayapuri, New Delhi
Vap: 306 07 306 07 306 07

D.W & Waste Tubings	2	
Sample Probe	1	
Power Cords	1	

Validation Team:

Name: Mr. Arun Yadav

Signature: 

Designation: DRSM

Date: 07-Jun-2022

C. List of Manuals, Certificates and Drawings

Meril Diagnostics Pvt. Ltd. provides the following with the instrument.

1. Operator's/User's manual for AutoQuant 200i.

D. Change Control Procedure

The instrument will not be altered, enhanced or substituted for another system until a formal change Control Authorization is approved from Meril Diagnostics Pvt. Limited.

E. Maintenance

The instrument listed within this document will placed under the control of purchasing / contracted institution with respect to proper maintenance procedures as detailed in the operations manual. The maintenance procedures will be filed separately.

A trained analyst using the manuals provided with the instrumentation can perform simple maintenance. Upon expiration of the warranty period Meril Diagnostics offers several level of maintenance agreements and performance testing services to assist you in maintaining GLP/GMP compliance. Containing your local representative and requesting the additional service Agreement can supply additional information.



T : +91 260 3052100

F : +91 260 3052125

www.meril.com

Meril Diagnostics Pvt. Ltd.
 Survey No. 13324
 Mukta, 1 Marg, Chala
 Vapi - 396195, Gujarat, India

F. Spare Parts

Meril Diagnostics Pvt. Ltd. strongly recommends the end user maintain a basic of consumable parts onsite to minimize down time due to minor failures. They have provided the list of such consumable parts.

For RRC, maintenance will be taken care as per contractual agreement.

Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature: 

Date: 07-Jun-2022

VI. INSTALLATION PROCEDURE:

(These have been performed at the time of original installation at the initial location)

1. Installation of Hardware

Follow the instructions mentioned in section 1 of the operator's manual for installation of the hardware.

2. Installation of Software

(Pre installed software is provided with the system.)

3. Installation of Printer

Follow the instructions mentioned in Setup Guide of HP color ink jet printer.



T : +91 260 3052100

F : +91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No 136/2 B, Ganga Park
Muktanand Marg, Chala
Vapi - 396191 Gujarat, India

Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support


Signature: 

Date: 07-Jun-2022

VII. COMMENTS

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature: 

Date: 07-Jun-2022



T : 91 960 3052100
F : 91 960 3052125
www.merilids.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135/2/B, Block 13/10,
Mustafabad Marg, Connaught
Place, New Delhi - 110029

VIII. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outlined in this Installation Qualification protocol, or exceptional conditions have identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Signature: 


Title: INSTALLATION QUALIFICATION

Date: 07-Jun-2022

Company: MERIL DIAGNOSTICS PVT. LIMITED

Customer Authorizations:

Name : 

Signature: 

Title :

Date: 07/Jun/2022



1. READ AND UNDERSTAND THE PROTOCOL
2. SIGNATURE OF THE REPRESENTATIVE
3. SIGNATURE OF THE CUSTOMER
4. DATE OF COMPLETION
5. DATE OF REVIEW
6. DATE OF RE-EVALUATION

AutoQuant 200i
Fully Automated Biochemistry Analyzer

OPERATIONAL QUALIFICATION

FOR

Delhi Clinical Laboratory - Haryana

OPERATIONAL QUALIFICATION#02

MARKETED BY:

MERIL DIAGNOSTICS PVT. LIMITED

MUKTANAND MARG, CHALA

VAPI - 396191

INDIA



T - 91 260 3052100
F - 91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 1234, Road No. 1234
Muktanand Marg, Chala
Vapi - 396191, Gujarat, India

TABLE OF CONTENT

Sr. No. CONTENTS

- I. Approval of the OQ procedure
- II. Instructions
- III. Scope
- IV. Operational Qualification
- V. Comments
- VI. System Certification



T : +91 260 3052100
F : +91 260 3052125
www.meril.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135 2 B, Block
Mustafai Marg, Conda
Vadgaon, Pune, India

I. APPROVAL OF THE OQ PROCEDURE:

Meril Diagnostics Pvt. Ltd. is responsible for the installation of the **Fully Automatic Biochemistry Analyzer, Merilyzer - AutoQuant 200i, Sr. No. AQ-050974** at **Delhi Clinical Laboratory - Haryana** as per the attached protocol.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Signature: 

Title: OPERATIONAL QUALIFICATION

Date: 07-Jun-2022

Company: MERIL DIAGNOSTICS PVT. LTD.

Validation Team from CPU:

Name: Mr. Rahul Saini

Signature: 


Title: DRSM – Technical & Scientific Support

Date: 07-Jun-2022

Department: MERIL DIAGNOSTICS PVT. LTD

Customer Authorizations:

Name: Amit

Signature: 

Title :

Date: 07/20/22



T +91 260 3052100
F +91 260 3052125
www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No 135/2-B, Block-2, H-15
Mustafabad Marg, Ghaziabad
U.P. - 201011, India

II. INSTRUCTIONS

1. An authorized Meril Diagnostics Pvt. Ltd. representative will check each module and enter the specific data as outlined in the appropriate Operational Qualification. Each result will be noted and dated.
2. Employees of (customer) Delhi Clinical Laboratory - Haryana will verify each result and sign in the last page. The members of the validation will carry this out.
3. All deviation from the acceptance criteria detailed in this document will be noted under COMMENTS section at the end of the OQ protocol. All resolution to such problems will also be noted in the COMMENTS section and must be resolved prior to issuance of a SYSTEM CERTIFICATION.
4. Any test data, which does not meet the specified acceptance criteria, will be submitted to the appropriate laboratory personnel for solution. All steps taken subsequently will be documented.

Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature: 

Date: 07-Jun-2022



T +91 260 3052100

F +91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 13/2018
Mukta Bahar Marg, Chhala
Vapi - 390015, Gujarat, India

II. SCOPE

This operational Qualification protocol will be performed on the **Fully Automatic Biochemistry Analyzer, Merilyzer AutoQuant 200i, Sr. No. AQ-050974** located at **Delhi Clinical Laboratory - Haryana**. This protocol will define the documentation that will be used to evaluate the instrument operation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that this instrument has been installed in accordance with the intended usage.

Trained, knowledgeable personnel will perform qualification studies.

Any exceptional condition encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Validation Team:

Name: Mr. Rahul Saini

Designation: DRSM - Technical & Scientific Support

Signature: 

Date: 07-Jun-2022



T - 91 260 3052100

F - 91 260 3052125

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 106 - B, Balarhuda
Mustafanand Marg, Chhata
Vapi - 396191, Gujarat, India

II. OPERATIONAL QUALIFICATION

A. Instrument Identification Verified Date

1. Model Name **AutoQuant 200i**
2. Serial Number **AQ-050974**

B. Following is the list of test performed and verified

Test No.	Test Name	Test Purpose	Verified Date
1.	Assay Programming	Programming a test as per the kit insert	07-Jun-2022
2.	Worklist Programming	In order to set up for a RUN	
3.	Initialize	To do the initialization of daily activity	
4.	Print	Printing Of Results	
5.	End of Work	For maintenance of the instrument	

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



T : +91 260 3052100
F : +91 260 3052125
www.meril.com

Meril Diagnostics Pvt. Ltd
Survey No 135 2/B, Bilalpur Has
Muktanand Marg, Chala
Vap. 386191 Gujarat, India

C. Operational Testing**Test I****Test Name: Assay Programming****Purpose: To programme the specified test****Procedure:**

1. Open AutoQuant 200i Operating Software.
2. Click on Parameter button
3. Enter test name, test code, define Method, Filter and other details menu.
4. Define Reference range in Reference Range window.
5. Define Calibrator in Calibration Window.
6. Input Control values in QC window.
7. Input Calculation formula in calculation item setting window.
8. Click on Save to save the protocol.

Validation Team:**Name: Mr. Arun Yadav****Designation: Sr. Exe. Application Support****Signature: Arun****Date: 07-Jun-2022**

T : +91 200 3052163

F : +91 200 3052123

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 15/3, Sector 10, D.P.
Mesa and Mang Circle
Vapi - 386 09, Gujarat, India

D. Creating & Processing a Worklist

Test II

Purpose: In order to set up for a worklist of same type of multiple tests to be run together.

Method:

1. Click on Schedule button.
2. Click on sample button.
3. Select test to be processed.
4. Input patient information.
5. Click on Add and Modify button,
6. Click ID++ to programme next sample,
7. Put samples on designated positions,
8. Click Save then Test button to start processing.
9. Refer Operators manual section 4 for more details.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



Tel: +91 260 3052100
Fax: +91 260 3052125
www.meril.com

Meril Diagnostics Pvt. Ltd.
Survey No. 12, A-1, Sector 17,
Gurgaon, Haryana, India
Surveys & Services Division

E. Initialization

Test III

Test Name: **Initialization**

Purpose: In order to check the proper working of the instrument

Method:

1. The machine performs a self-test at the time of initialization to check if all the mechanics of the machine are working OK.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: Arun

Date: 07-Jun-2022



For more information
Visit our website
www.merildiagnostics.com
Meril Diagnostics
110, Sector 15, Gurgaon
Haryana - 122001, India



G. OPERATIONAL PROCEDURE

A. Certificate of training

1. **Technician Training :** This certifies that the technicians listed below have received basic user training in the following categories for the system described in this installation qualification.

Mr. Arun Yadav who is certified by Meril Diagnostics Pvt. Ltd. Has conducted the training.

Sr. No.	Training Program	Initials	Date
1.	Instrument Setup	OK	07-Jun-2022
2.	System operation	OK	07-Jun-2022
3.	Basic troubleshooting	OK	07-Jun-2022
4.	Routine Maintenance	OK	07-Jun-2022

2. **Operator Training:** The users responsible for the operation of this instrument will be trained in the proper usage of the system. Training will focus on the basic operation and maintenance of the system. The training of the operators will be documented and the training records will be filled as indicated below:

Sr. No.	Operators	Location	Initials	Date
1.		Delhi Clinical Laboratory - Haryana	OK	07-Jun-2022
2.			OK	

Validation Team:

Name: Mr. Arun Yadav Designation: Sr. Exe. Application Support

Signature: Date: 07-Jun-2022




T - 91 260 3082100
F - 91 260 3082121
www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 123/14, Sector 14, Gurgaon
Haryana - 122002, India
Aapko 360° se dekhne do.

Cardiovascular

Orthopedic




V. COMMENTS

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



Meril Diagnostics
S. No. 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94, 95, 96, 97, 98, 99, 100
www.meril.co.in

Meril Diagnostics
S. No. 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94, 95, 96, 97, 98, 99, 100
Mumbai, Maharashtra
Pin - 400 001, Gujarat, India

VI. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outlined in this Operational Qualification protocol, or exceptional conditions have identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Signature: 

Title: OPERATIONAL QUALIFICATION

Date: 07-Jun-2022

Company: MERIL DIAGNOSTICS PVT. LIMITED

Customer Authorizations:

Name: Amit

Signature: 

Title :

Date: 

Site :



Tel: +91 260 3082100

Fax: +91 260 380125

Meril Diagnostics Pvt. Ltd.

Sector No. 22, Phase 1

MCC Road, Vasant Vihar

New Delhi - 110021, India



AutoQuant 200i
Fully Automated Biochemistry Analyzer

PERFORMANCE QUALIFICATION

FOR

Delhi Clinical.Laboratory - Haryana

PERFORMANCE QUALIFICATION#01

MARKETED BY:

MERIL DIAGNOSTICS PVT. LIMITED

MUKTANAND MARG, CHALA

VAPI - 396191

INDIA



T - 91 260 3052100

F - 91 260 3052125

www.meril-di.com

Meril Diagnostics Pvt. Ltd.
Survey No. 1/57-B (Block) H-15
Muktanand Marg, Chala
Vapi - 396191, Gujarat, India

TABLE OF CONTENTS

Sr. No. CONTENTS

- I. Approval of the PQ procedure
- II. Instructions
- III. Scope
- IV. Performance Qualification
- V. Comments
- VI. System Certification



T + 91 260 3052100
F + 91 260 3052125
www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No 135 2/B, Blakha, Huda
Muktanand Marg, Chata,
Vapi - 396191, Gujarat, India

I. APPROVAL OF THE PQ PROCEDURE:

Meril Diagnostics Pvt. Ltd. is responsible for the installation of the **Fully Automatic Biochemistry Analyzer, Merilyzer - AutoQuant 200i, Sr. No. AQ-050974** at Delhi Clinical Laboratory - Haryana as per the attached protocol.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Signature: 

Title: OPERATIONAL QUALIFICATION

Date: 07-Jun-2022

Company: MERIL DIAGNOSTICS PVT. LTD.

Validation Team from CPU:

Name: Mr. Rahul Saini

Signature: 

Title: DRSM - Technical & Scientific Support

Date: 07-Jun-2022

Department: MERIL DIAGNOSTICS PVT. LTD

Customer Authorizations:

Name : 

Signature: 

Title :

Date: 07/24/22

Site :



Tel : +91 760 3052100
F : +91 760 3052125
www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 13573, Bhubar H. U.S.
Mustafand Marg, Chala
Vap. - 396191, Gujarat, India

II. INSTRUCTIONS

1. An authorized Meril Diagnostics Pvt. Ltd. representative will check for the performance of the system and enter the specific data as outlined in the appropriate Performance Qualification.
2. Performance checks on a regular basis described in the Further Performance Checks (vide-infra) will be responsibility of the customer's personnel.
3. Employees of (customer) Delhi Clinical Laboratory - Haryana will verify each result and sign in the last page. The members of the validation will carry this out.
4. All deviation from the acceptance criteria detailed in this document will be noted under COMMENTS section at the end of PQ section. All resolution to such problems will also be noted in the COMMENTS section and must be resolved prior to issuance of a SYSTEM CERTIFICATION.
5. Any test that does not meet the specified acceptance criteria will be submitted to the appropriate laboratory personnel for solution. All steps taken subsequently will be documented.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022

T: +91 260 3052103
F: +91 260 3052125
www.meriltd.com



Meril Diagnostics Pvt. Ltd.
Survey No. 12/1, B-1, Sector 17, Gurgaon
Mustafa Ali Marg, Connaught Place
Vap: 80659, Gurgaon, Haryana

III. SCOPE

This Performance Qualification protocol will be performed on the **Fully Automatic Biochemistry Analyzer, Merilyzer AutoQuant 200i, Sr. No. AQ-050974** located at **Delhi Clinical Laboratory - Haryana**. This protocol will define the documentation that will be used to evaluate the instrument operation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified is performing in accordance with the intended usage.

Trained, knowledgeable personnel will perform qualification studies.

Any exceptional condition encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



T : +91 200 3052100
F : +91 200 3052125
www.meril.de.com

Meril Diagnostics Pvt. Ltd.
Sector No. 12, Phase - II, Gurgaon
Haryana - 122002, India
Vap: 39679, Sector 12, Gurgaon

IV. PERFORMANCE QUALIFICATION

A. Instrument Identification Verified Date

1. Model Name AutoQuant 200i.
2. Serial Number AQ-050974

B. Following is the list of test performed and verified

Test No. Test Name Test Purpose Verified Date

- 01: SELF TEST : In order to check the proper working of the instrument.
- 02: DAILY MAINTENANCE : To eliminate unexpected errors.
- 03: RUN A TEST : To check the instrument performance.
04. CARRY OVER STUDY : To check carry over possibilities.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022

TEL: +91 30 30 2100
TEL: +91 30 30 2125
www.meril.com



Meril Diagnostics Pvt. Ltd.
Survey No. 13/1, B. Road, Sector 13, DDA
Mustafabad Marg, Okhla
New Delhi - 110021, India



C. Performance Testing

Test I

Test Name: Self-Test

Purpose: In order to check the proper working of the instrument

Method: Switch ON the machine and Initialize the machine.

Acceptance Criteria: The machine passes Self Test and comes to main menu screen.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



Tel: +91 330 3902100
Fax: +91 330 3902125
www.meril.com

Meril Diagnostics Pvt. Ltd.
Survey No. 125/33, Sector 14, Phase 1
Mukherjee Marg, Cuttack
Vap: 396191, Gujarat, India

CARRY OVER CHECK :

Validation of Glucose/Urea/SGOT test run:

Replicates processed : 20

Number of Test & Samples processed : 10

Carry over found during performance check : YES / NO

(Data Sheet Attached for reference)

PARAMETER: PASS / FAIL


During sample analysis run control samples with each test run. Validate test run as per the criteria stated in the respective kit inserts.

Accept the sample results only on meeting of the validation criteria.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



Tel: +91 260 80 2100
Fax: +91 260 80 2174

www.meril.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135/2B, Bhubaneswar H.O.
Mustafabad Marg, Cuttack
Pin - 753019, Odisha, India

Validation of Instrument Performance :

Validation Criteria	Result obtained	Criteria pass/fail	Verified by
Glucose Test	Ok.	Pass	
Urea Test	Ok	Pass	
SGOT Test	Ok	Pass	

PARAMETER: PASS / FAIL


During sample analysis run control samples with each test run. Validate test run as per the criteria stated in the respective kit inserts.

Accept the sample results only on meeting of the validation criteria.

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



T : 91 260 3052100
 F : 91 260 3052125
 www.meril.co.in

Meril Diagnostics Pvt. Ltd.
 Service Network
 Plot No. 10, Sector 10, Gurgaon
 Haryana - 122002

V. COMMENTS

Validation Team:

Name: Mr. Arun Yadav

Designation: Sr. Exe. Application Support

Signature: 

Date: 07-Jun-2022



T - 91 260 3052100
F - 91 260 3052125
www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135 2d, Rajahmundry
Mustanser Medical Challah
Vap - 390191



VI. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outlined in this Operational Qualification protocol, or exceptional conditions have identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.

Protocol Performed By: Meril Diagnostics Pvt. Ltd. Representative

Name: Mr. Arun Yadav

Signature:

Title: PERFORMANCE QUALIFICATION

Date: 07-Jun-2022

Company: MERIL DIAGNOSTICS PVT. LIMITED

Customer Authorizations:

Name :

Signature:

Title :

Date: 07/06/22

Site :



T : 91 50 3062100
F : 91 50 3062101

www.merilife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 107, H. S. Road, 107/108
Mustafabad Marg, Chhata
Vap. 396191 Gujarat, India

Qc results

From: 07/20/2022 To: 07/20/2022 Concentration: Normal

No.	Test name	QC lot	Result	Run date	Flag	QC range
	ALB	0522206	3.79	07/20/2022::07:42:14	Normal	3.29-4.01
	UREA	0522206	43.22	07/20/2022::07:43:00	Normal	37.35-53.85
	CREAT	0522206	1.22	07/20/2022::07:44:20	Normal	1.04-1.46
	TBIL	0522206	1.61	07/20/2022::07:46:29	Normal	1.20-2.24
	UA	0522206	5.85	07/20/2022::07:46:29	Normal	4.65-6.55
	TP	0522206	5.68	07/20/2022::07:49:41	Normal	5.13-6.27
	GLUC	0522206	80.58	07/20/2022::07:49:51	<-1SD	77.60-109.40
	CHOL	0522206	133.41	07/20/2022::07:50:16	Normal	108.02-152.28
	TG	0522206	110.33	07/20/2022::07:50:54	Normal	91.50-128.70

✓

Qc results

From: 07/20/2022 To: 07/20/2022 Concentration: High

No.	Test name	QC lot	Result	Run date	Flag	QC range
	ALB	MB052204	4.19	07/20/2022::09:16:56	Normal	3.69-4.51
	UREA	MB052204	118.99	07/20/2022::09:17:58	<-1SD	112.05-157.95
	CREAT	MB052204	3.43	07/20/2022::09:19:16	Normal	2.66-3.74
	TBIL	MB052204	4.76	07/20/2022::09:20:47	Normal	3.43-6.37
	UA	MB052204	8.59	07/20/2022::09:21:00	Normal	6.85-9.65
	GLUC	MB052204	217.59	07/20/2022::09:24:34	<-1SD	199.20-280.80
	TP	MB052204	6.50	07/20/2022::09:24:34	Normal	5.76-7.04
	CHOL	MB052204	155.86	07/20/2022::09:24:54	Normal	132.40-187.60
	TG	MB052204	169.89	07/20/2022::09:25:37	Normal	146.89-207.11

