

05-08-2022 11:19:34

23.8546386

73.7068294





AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 09073F	Field	Mech.	Group	Mass	Page No.	1 of 1
SRF No & Date	220524.4 DT - 24 / 05 / 2022			U _L R No.	CC2646220000 09073F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Medicentre Sonography & Clinical Lab 01 Behind LIC Building, Opp. RTO Hospital Road, Dungarpur (Rj.)			Calibration Date	24.05.2022		
Address				Cali. Due On (As Sugg. By Cust.)	23.05.2023		
				Certificate Issue Dt.	27.05.2022		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	0.1 µl	Precica / -----	-----

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/252/JAN/22	06/03/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	100	99.98186	0.01814	0.25
2	500	499.93544	0.06456	0.25
3	1000	999.90405	0.09595	0.25

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

Asish
Calibrated by



Approved by
Anand Chaurasia
Anand Chaurasia
(CEO)