



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2022



Lab Name	REDCLIFFE LIFETECH PVT LTD	Lab No	15550
Constituent Group	HbA1c	Date of Result Entered :	18/07/2022
PT item	Lyophilized Whole blood	Date of Report Published :	05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Any Analyser	279	6.35	11.65	0.74	6.7 %	0.47	0.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

REDCLIFFE LIFETECH PVT LTD
C/O SERA PATHOLOGY, 12/475, ULTRA TOWER, MICROBERTGANJ, NEAR SANI DEV MANDIR, CHUNNIGANJ
KANPUR
UTTAR PRADESH 208001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****