



# AN LABORATORIES

## CALIBRATION & TESTING LABORATORIES

303, B-56, Kapil CHS, Gokuldham, Goregaon (East), Mumbai - 400 063.  
 Mob.: 7021536885 / 9969541543 E-mail : sales.anlaboratories@gmail.com

### Calibration Certificate

Centrifuge

Certificate No: AN/22020018

Service Requested by:		M/s Asavlee Dr. Aparna's Pathology Laboratory	Request Date:	05.02.2022
			Received on:	07.02.2022
			Physical Condition:	Good
<b>Details of Instrument/Unit Under Calibration (UUC)</b>				
Instrument Name:	Centrifuge	Calibration Date:	07.02.2022	
Manufacturer:	Remi	Calibration Due Date:	06.02.2023	
Model:	R303	Calibrated at:	Site	
Serial No.:	ZCKN-35540			
Id. No.:	NA			
Location:	NA	Equipment :	NA	
<b>Ambient Conditions</b>				
Temperature	25.9 ± 2 °C	Relative Humidity	58 %RH ± 5 %RH	
<b>Calibration Specifications</b>				
Parameters	Range	Calibration Points	Accuracy	Resolution
Speed	3500 RPM	NA	Not Specified	NA
<b>Reference Standards Used</b>				
Equipment Type	Serial No./Id. No.	Uncertainty	Certificate No.	Cal. Date
Photo / Contact Tachometer	Q633279/ AN-001	± 3.1 RPM	QATM-21060493	21-06-2021
				Due Date
				20-06-2022
<b>Calibration Procedure:</b> Centrifuge is calibrated by comparison with standard Photo / Contact Tachometer				
<b>Observations</b>				
UUC Reading	Standard Readings	Uncertainty		
RPM	RPM	RPM		
Step 1	1680	± 8		
Step 2	2477			
Step 3	3105			
Step 4	3467			
<p>Note:          UUC: Unit under calibration          Uncertainty has been calculated for a coverage factor k=2 corresponding to approximately 95.5% Confidence Level          The Standard maintained are traceable to National/ International Standard through accredited Laboratories.          The observations reported represent values at the time of the measurements and under the stated condition. They do not convey any long term stability information.</p>				
		Calibrated By:	MA 08-feb-2022	
		Name:	Manish Aple	
		Reviewed By:	Bhale 08/02/2022	
		Name:	Jayashree Bhale	
END OF REPORT				
Page 01 of 01				

⊗ Typo error

Bhale.

08/02/2022