



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2022



Lab Name: DR APARNA'S PATHOLOGY LABORATORY Lab No: 15380
 Constituent Group: Chemistry II Date of Result Entered: 20/07/2022
 PT item: Lyophilized Serum Date of Report Published: 05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	Dry Chemistry	Fuji Dry Chemistry series	21	312.65	7.05	22.03	309 mg/dL	-0.17	9.62
2	UREA II	Dry Chemistry	Fuji Dry Chemistry series	29	59.99	7.64	4.58	59.2 mg/dL	-0.17	1.70
3	CREATININE II	Dry Chemistry	Fuji Dry Chemistry series	18	3.42	7.87	0.27	3.5 mg/dL	0.30	0.13
4	T.BILIRUBIN II	Dry Chemistry	Fuji Dry Chemistry series	35	1.76	16.85	0.30	1.6 mg/dL	-0.54	0.10
5	T-PROTEIN II	Dry Chemistry	Fuji Dry Chemistry series	33	5.68	8.05	0.46	5.3 g/dL	-0.83	0.16
6	ALBUMIN II	Dry Chemistry	Fuji Dry Chemistry series	29	3.33	5.50	0.18	3.2 g/dL	-0.71	0.07
7	URIC ACID II	Dry Chemistry	Fuji Dry Chemistry series	35	7.88	10.63	0.84	7.9 mg/dL	0.02	0.28
8	CHOLESTROL II	Dry Chemistry	Fuji Dry Chemistry series	28	119.29	11.58	13.82	121 mg/dL	0.12	5.22
9	TRIGLYCERIDE II	Dry Chemistry	Fuji Dry Chemistry series	29	299.60	12.87	38.57	327 mg/dL	0.71	14.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DR APARNA'S PATHOLOGY LABORATORY
SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI
MUMBAI
MAHARASHTRA400093

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: [0416-2283102](tel:0416-2283102)

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2022



Lab Name **DR APARNA'S PATHOLOGY LABORATORY** Lab No **15380**
Constituent Group **Chemistry II** Date of Result Entered : **19/01/2022**
PT item **Lyophilized Serum** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	Dry Chemistry	Fuji Dry Chemistry series	17	121.17	5.25	6.36	125 mg/dl	0.60	3.09
2	UREA II	Dry Chemistry	Fuji Dry Chemistry series	23	47.65	7.12	3.39	50.72 mg/dl	0.91	1.41
3	CREATININE II	Dry Chemistry	Fuji Dry Chemistry series	20	1.81	8.39	0.15	1.73 mg/dl	-0.53	0.07
4	T.BILIRUBIN II	Dry Chemistry	Fuji Dry Chemistry series	23	2.22	11.75	0.26	2.2 mg/dl	-0.08	0.11
5	T-PROTEIN II	Dry Chemistry	Fuji Dry Chemistry series	25	4.89	6.85	0.34	5.1 g/dl	0.63	0.13
6	ALBUMIN II	Dry Chemistry	Fuji Dry Chemistry series	25	3.08	6.00	0.18	3.2 g/dl	0.65	0.07
7	URIC ACID II	Dry Chemistry	Fuji Dry Chemistry series	29	6.98	6.16	0.43	7 mg/dl	0.05	0.16
8	CHOLESTROL II	Dry Chemistry	Fuji Dry Chemistry series	16	101.57	7.69	7.81	105 mg/dl	0.44	3.91
9	TRIGLYCERIDE II	Dry Chemistry	Fuji Dry Chemistry series	21	129.76	9.36	12.15	140 mg/dl	0.84	5.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DR APARNA'S PATHOLOGY LABORATORY
SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI
MUMBAI
MAHARASHTRA400093

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******