



**CHRISTIAN MEDICAL COLLEGE**  
**DEPARTMENT OF CLINICAL BIOCHEMISTRY**  
**CMC EXTERNAL QUALITY ASSURANCE SCHEME**  
**MONTHLY SUMMARY REPORT - JULY 2022**

Lab Name	DR APARNA'S PATHOLOGY LABORATORY	Lab No	15380
Constituent Group	Urine (Microalbumin & Protein)	Date of Result Entered :	20/07/2022
PT item	Lyophilized Urine	Date of Report Published :	05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	UR. PROTEIN	Ur. Pyrogallol	Any Analyser (Automation / Semi Automation )	71	24.62	21.83	5.37	18.2 mg/dL	-1.19	1.28
2	UR. MICROALBUMIN	Ur. Immuno Turbidimetric	Any Analyser (Automation / Semi Automation )	109	112.93	19.33	21.83	57 mg/L	-2.56	4.18

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**DR APARNA'S PATHOLOGY LABORATORY**  
**SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI**  
**MUMBAI**  
**MAHARASHTRA400093**

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: [0416-2283102](tel:0416-2283102)

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**  
**Data in CMC EQAS reports is confidential**  
**CMC EQAS does not sub contract any components**  
**\*\*\*\*\* End of Report \*\*\*\*\***



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2022

Lab Name **DR APARNA'S PATHOLOGY LABORATORY** Lab No **15380**  
 Constituent Group **Urine (Microalbumin & Protein)** Date of Result Entered : **20/01/2022**  
 PT item **Lyophilized Urine** Date of Report Published : **07/02/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	UR. PROTEIN	Ur. Pyrogallol	Any Analyser (Automation / Semi Automation )	45	37.46	19.36	7.25	38.2 mg/dL	0.10	2.16
2	UR. MICROALBUMIN	Ur. Immuno Turbidimetric	Any Analyser (Automation / Semi Automation )	57	207.24	13.61	28.20	130.4 mg/L	-2.72	7.47

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

**DR APARNA'S PATHOLOGY LABORATORY**  
**SHOP 12, DHEERAJ GARDEN CHS, OPP ALN PARK, POONAM NAGAR, ANDHARI**  
**MUMBAI**  
**MAHARASHTRA400093**

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

