

Calibration Certificate

This is certified that following instrument installed in Green Labs & Diagnostics, No.3/199, 1st Main Road, Perumbakkam, Near MSK Clinic, Chennai – 600100.

Instrument Type	Clinical Chemistry Analyzer			
Instrument Model	LABMATE			
Serial no	ES1001701C087			
Done by	KISHORE KUMAR .P			
Date	Done on	19 Mar 2022	Expiry on	18 Mar 2023

Calibration Procedure

1. **Cleaning & Decontamination:**

Cleaned Main unit removed Dust & Rust. Decontamination had completed with preparation of 5% Sodium hypo chloride, applied & Cleaned thoroughly in instrument – Yes

2. **Lubrication & Rust Protection:**

Lubrication had done with silicon oil & Rust protection - Yes

3. **Cleaning Parts:**

- Flow cell - ok
- Main board assembly - ok
- Incubator - ok
- Aspiration tube - ok

4. **AD Calibration:**

Filter	AD	Gain	Offset
340	36229	255	227
405	54052	140	118
450	54018	205	179
505	53188	68	58
546	52982	52	46
578	53794	54	48
630	53636	34	29
670	52629	24	22

Trivitron Healthcare Pvt. Ltd.

15, IV Street, Abhiramapuram, Chennai - 600 018, India.
CIN: U85110TN1998PTC040515
Ph: +91 - 44 - 4347 7800 / 2498 5050. Fax: +91 - 44 - 2498 5757 / 2467 / 2782
E-mail: corporate@trivitron.com, Website: www.trivitron.com

BRANCH OFFICES: Mumbai (+91-22-2768 2211) | Delhi (+91-11-4135 4650) | Bengaluru (+91-80-2656 5491) | Kolkata | Pune
INTERNATIONAL: USA | Finland | Turkey | Africa | China | Dubai | Abu Dhabi | Singapore



005

2022/8/9 21:14