


# STATUS PASSED CALIBRATION REPORT

**DEVICE ID:**  
**METHOD ID:** VV1000  
**ENVIRONMENTAL FACTORS**  
**TEMP:** 20°C  
**AIR DENSITY:** 0.001200

**DESCRIPTION:** VAR VOL 100-1000µl  
**DATE:** 29/05/2019  
**REL.HUMIDITY:** 50%  
**CUBIC EXP:** 0.0000cm<sup>3</sup>/degc.

SAMPLE			CALIBRATION STATISTICS						
IMPRECISION CV%			INACCURACY E% IMPRECISION CV%						
VOL	NO	Wt	VOL	VOL	MEAN	SD	ACTUAL	TARGET	STATUS
100.0	1	100.00	100.30	100.0	100.29	0.008	1.00	0.80	passed<0.30
100.0	2	199.99	100.30	500.0	499.14	1.056	0.17	0.60	passed<0.20
100.0	3	299.99	100.30	1000.0	1000.96	1.861	0.09	0.50	passed<0.20
500.0	1	497.80	499.24						
500.0	2	996.50	500.15						
500.0	3	1493.10	498.04						
1000.0	1	1000.00	1002.90						
1000.0	2	1997.90	1000.79						
1000.0	3	2994.20	999.19						

All Pipettes are calibrated at 20°C with degassed double distilled water conforming to DIN12650

Checked By: 

Manager (Quality Control) 