



CERTIFICATE OF CALIBRATION

Certificate No : CC-22-23-S-530/06		Page. No: 1 of 1
Customer Name & Address	ULR No.	CC272522000004778F
M/S, Life Care Diagnostics	SRF No.	VCI-22-23-S-530
11, Vaithiyanatha Iyyer,	Date of Receipt	05.07.2022
1st Street, Madurai - 625 020.	Date of Calibration	05.07.2022
	Due Date of Calibration	04.07.2023
	Date of Issue	06.07.2022

Description	Deep Freezer	Deep Freezer Make	
Range	-20°C	Model	HF-155W
Resolution	0.1°C	Condition of UUC	Good
Serial Number	B30KE5E1X00BKKCW0014	Instrument Location	Molecular Biology Lab
ID Number	LCD/MDU/MOL/EQP/03	Cal. At	On Site

	Environmental Condition	Calibration Method Used		
Temperature (°C)	30±5	National / International Standard	DKD-R-5-7	
Humidity (%RH)	35 - 75	Cal Procedure No	VCI-L03-WI-THER-005	

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-005 & VCI-ET-DMM-001	CC-22-23-L-015/01	Tempsens	National Standards	08.01.2023

Results of Calibration					
Sl. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Measurement Uncertainty ± (°C)	
01	-19.85	-20.0	-0.15	0.36	

Remarks

- 1. UUC is Defined as the Unit Under Calibration.
- 2. This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/part without approval of the Venus Calibration and Instruments.
- 3. The reported results are valid at the time of under stated condition of measurement.
- 4. Calibration of the UUC are traceable to National / International Standards.
- 5. The recalibration interval shall be decided by the user.
- 6. Nabl-133 guidelines are adopted for use of NABL symbol.
- 7. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k = 2.
- 8. Temperature Chart: ITS-90

Checked & Issued by:

M. Manju

(Engineer - Calibration)



Authorised by:

C.Dhinakaran

(Technical / Quality Manager)