

Result Reprint

Report Type : Controls

23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval(3SD)
1	S062125B	ERBA NORM	GPTHL	35.2 U/L		17278	23-Dec-2021 11:24:56	33.600	2.240	26.88 - 40.32
2	S062125B	ERBA NORM	GOTHL	32.7 U/L		17279	23-Dec-2021 11:27:02	34.400	2.290	27.53 - 41.27
3	S062125B	ERBA NORM	ALPU	130 U/L		17280	23-Dec-2021 11:29:08	127.750	8.520	102.19 - 153.31
4	S062125B	ERBA NORM	GGT	39.8 U/L		17281	23-Dec-2021 11:31:15	41.970	2.800	33.57 - 50.37
5	S062125B	ERBA NORM	PHOS	4.42 mg/dL		17282	23-Dec-2021 11:33:21	4.730	0.240	4.01 - 5.45
6	S062125B	ERBA NORM	BIDD	0.58 mg/dL		17283	23-Dec-2021 11:35:27	0.580	0.050	0.43 - 0.73
7	S062125B	ERBA NORM	BILT	1.73 mg/dL	+1SD	17284	23-Dec-2021 11:37:34	1.510	0.130	1.12 - 1.9
8	S062125B	ERBA NORM	UREA	32.9 mg/dL		17285	23-Dec-2021 11:47:29	33.970	1.700	28.87 - 39.07
9	S062125B	ERBA NORM	CRENZ	1.17 mg/dL		17286	23-Dec-2021 11:49:36	1.140	0.070	0.93 - 1.35
10	S062125B	ERBA NORM	GLU	94.7 mg/dL		17287	23-Dec-2021 11:51:42	95.720	4.780	81.38 - 110.06
11	S062125B	ERBA NORM	TRIG	120.9 mg/dL	+1SD	17288	23-Dec-2021 11:53:49	113.350	5.670	96.34 - 130.36
12	S062125B	ERBA NORM	CHOL	136 mg/dL		17289	23-Dec-2021 11:55:55	137.570	6.880	116.93 - 158.21
13	S062125B	ERBA NORM	HDLC	41.3 mg/dL		17290	23-Dec-2021 11:58:01	42.860	2.860	34.28 - 51.44
14	S062125B	ERBA NORM	UA	6.5 mg/dL	+1SD	17291	23-Dec-2021 12:00:07	6.080	0.300	5.18 - 6.98
15	S062125B	ERBA NORM	CA	8.4 mg/dL		17292	23-Dec-2021 12:02:11	8.090	0.400	6.89 - 9.29
16	S062125B	ERBA NORM	ALBD	3.86 mg/dL	+1SD	17561	23-Dec-2021 12:04:20	3.570	0.180	3.03 - 4.11
17	S062125B	ERBA NORM	PRO	6.38 g/dL		17562	23-Dec-2021 12:10:57	6.240	0.310	5.31 - 7.17

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
1	12345	BLANK	SGPTD	-0.0004	-	1	23-Dec-2021 13:23:07
2	12345	BLANK	SGPTD	-0.0005	-	2	23-Dec-2021 13:23:25
3	12345	BLANK	SGPTD	-0.0001	-	3	23-Dec-2021 13:23:43
4	2105079B	MULTICAL	SGPTD	-0.0378	-	4	23-Dec-2021 13:24:01
5	2105079B	MULTICAL	SGPTD	-0.0374	-	5	23-Dec-2021 13:24:19
6	2105079B	MULTICAL	SGPTD	-0.0374	-	6	23-Dec-2021 13:24:38
7	12345	BLANK	SGOTD	-0.0007	-	8	23-Dec-2021 13:25:14
8	12345	BLANK	SGOTD	-0.0008	-	9	23-Dec-2021 13:25:32
9	12345	BLANK	SGOTD	-0.0006	-	10	23-Dec-2021 13:25:50
10	2105079B	MULTICAL	SGOTD	-0.0396	-	11	23-Dec-2021 13:26:08
11	2105079B	MULTICAL	SGOTD	-0.0392	-	12	23-Dec-2021 13:26:26
12	2105079B	MULTICAL	SGOTD	-0.0393	-	13	23-Dec-2021 13:26:44
13	12345	BLANK	ALPU	0.0005	-	15	23-Dec-2021 13:27:20
14	12345	BLANK	ALPU	0.0005	-	16	23-Dec-2021 13:27:38
15	12345	BLANK	ALPU	0.0005	-	17	23-Dec-2021 13:27:56
16	2105079B	MULTICAL	ALPU	0.0805	-	18	23-Dec-2021 13:28:14
17	2105079B	MULTICAL	ALPU	0.0830	-	19	23-Dec-2021 13:28:32

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
18	2105079B	MULTICAL	ALPU	0.0835 -		20	23-Dec-2021 13:28:50
19	12345	BLANK	GGT	0.0022 -		22	23-Dec-2021 13:29:26
20	12345	BLANK	GGT	0.0014 -		23	23-Dec-2021 13:29:44
21	12345	BLANK	GGT	0.0016 -		24	23-Dec-2021 13:30:02
22	2105079B	MULTICAL	GGT	0.0706 -		25	23-Dec-2021 13:30:21
23	2105079B	MULTICAL	GGT	0.0707 -		26	23-Dec-2021 13:30:39
24	2105079B	MULTICAL	GGT	0.0710 -		27	23-Dec-2021 13:30:57
25	12345	BLANK	PHOS	0.2724 -		29	23-Dec-2021 13:31:33
26	12345	BLANK	PHOS	0.2656 -		30	23-Dec-2021 13:31:51
27	12345	BLANK	PHOS	0.2650 -		31	23-Dec-2021 13:32:09
28	2105079B	MULTICAL	PHOS	0.5658 -		32	23-Dec-2021 13:32:27
29	2105079B	MULTICAL	PHOS	0.5812 -		33	23-Dec-2021 13:32:45
30	2105079B	MULTICAL	PHOS	0.5796 -		34	23-Dec-2021 13:33:03
31	12345	BLANK	BIDD	0.0035 -		37	23-Dec-2021 13:33:57
32	12345	BLANK	BIDD	0.0032 -		38	23-Dec-2021 13:34:15
33	2105079B	MULTICAL	BIDD	0.1183 -		39	23-Dec-2021 13:34:33
34	2105079B	MULTICAL	BIDD	0.1186 -		40	23-Dec-2021 13:34:51

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
35	2105079B	MULTICAL	BIDD	0.1192 -		41	23-Dec-2021 13:35:09
36	12345	BLANK	BIT	0.0037 -		43	23-Dec-2021 13:35:45
37	12345	BLANK	BIT	0.0032 -		44	23-Dec-2021 13:36:03
38	12345	BLANK	BIT	0.0027 -		45	23-Dec-2021 13:36:22
39	2105079B	MULTICAL	BIT	0.2248 -		46	23-Dec-2021 13:36:40
40	2105079B	MULTICAL	BIT	0.2241 -		47	23-Dec-2021 13:36:58
41	2105079B	MULTICAL	BIT	0.2243 -		48	23-Dec-2021 13:37:16
42	12345	BLANK	UREA	-0.0022 -		50	23-Dec-2021 13:37:52
43	2105079B	MULTICAL	UREA	-0.1295 -		54	23-Dec-2021 13:46:53
44	2105079B	MULTICAL	UREA	-0.1291 -		55	23-Dec-2021 13:47:11
45	12345	BLANK	GLU	0.0154 -		57	23-Dec-2021 13:47:48
46	12345	BLANK	GLU	0.0156 -		58	23-Dec-2021 13:48:06
47	12345	BLANK	GLU	0.0160 -		59	23-Dec-2021 13:48:24
48	2105079B	MULTICAL	GLU	0.7153 -		60	23-Dec-2021 13:48:42
49	2105079B	MULTICAL	GLU	0.7029 -		61	23-Dec-2021 13:49:00
50	2105079B	MULTICAL	GLU	0.7087 -		62	23-Dec-2021 13:49:18
51	12345	BLANK	TRIG	0.0110 -		64	23-Dec-2021 13:49:54

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
52	12345	BLANK	TRIG	0.0105 -		65	23-Dec-2021 13:50:12
53	12345	BLANK	TRIG	0.0105 -		66	23-Dec-2021 13:50:30
54	2105079B	MULTICAL	TRIG	0.2671 -		67	23-Dec-2021 13:50:48
55	2105079B	MULTICAL	TRIG	0.2693 -		68	23-Dec-2021 13:51:06
56	2105079B	MULTICAL	TRIG	0.2673 -		69	23-Dec-2021 13:51:24
57	12345	BLANK	CHOL	0.0406 -		71	23-Dec-2021 13:52:00
58	12345	BLANK	CHOL	0.0403 -		72	23-Dec-2021 13:52:18
59	12345	BLANK	CHOL	0.0406 -		73	23-Dec-2021 13:52:36
60	2105079B	MULTICAL	CHOL	0.3037 -		74	23-Dec-2021 13:52:54
61	2105079B	MULTICAL	CHOL	0.3189 -		75	23-Dec-2021 13:53:12
62	2105079B	MULTICAL	CHOL	0.3121 -		76	23-Dec-2021 13:53:31
63	12345	BLANK	HDLC	0.0051 -		78	23-Dec-2021 13:54:07
64	12345	BLANK	HDLC	0.0050 -		79	23-Dec-2021 13:54:25
65	12345	BLANK	HDLC	0.0052 -		80	23-Dec-2021 13:54:43
66	B082122	HDL CAL	HDLC	0.1170 -		81	23-Dec-2021 13:55:01
67	B082122	HDL CAL	HDLC	0.1212 -		82	23-Dec-2021 13:55:19
68	B082122	HDL CAL	HDLC	0.1218 -		83	23-Dec-2021 13:55:37

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
69	12345	BLANK	UA	0.0042 -		85	23-Dec-2021 13:56:13
70	12345	BLANK	UA	0.0049 -		86	23-Dec-2021 13:56:31
71	12345	BLANK	UA	0.0041 -		87	23-Dec-2021 13:56:49
72	2105079B	MULTICAL	UA	0.0868 -		88	23-Dec-2021 13:57:07
73	2105079B	MULTICAL	UA	0.0860 -		89	23-Dec-2021 13:57:25
74	2105079B	MULTICAL	UA	0.0870 -		90	23-Dec-2021 13:57:43
75	12345	BLANK	CA	0.2025 -		92	23-Dec-2021 13:58:19
76	12345	BLANK	CA	0.1983 -		93	23-Dec-2021 13:58:37
77	12345	BLANK	CA	0.2021 -		94	23-Dec-2021 13:58:55
78	2105079B	MULTICAL	CA	0.5904 -		95	23-Dec-2021 13:59:14
79	2105079B	MULTICAL	CA	0.5709 -		96	23-Dec-2021 13:59:32
80	2105079B	MULTICAL	CA	0.5926 -		97	23-Dec-2021 13:59:50
81	12345	BLANK	ALBD	0.0773 -		99	23-Dec-2021 14:00:26
82	12345	BLANK	ALBD	0.0758 -		100	23-Dec-2021 14:00:44
83	12345	BLANK	ALBD	0.0769 -		101	23-Dec-2021 14:01:02
84	2105079B	MULTICAL	ALBD	0.7727 -		102	23-Dec-2021 14:01:20
85	2105079B	MULTICAL	ALBD	0.7739 -		103	23-Dec-2021 14:01:38

Result Reprint

Report Type : Calibration

08-Oct-2021 To 23-Dec-2021

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
86	2105079B	MULTICAL	ALBD	0.7796 -		104	23-Dec-2021 14:01:56
87	12345	BLANK	PRO	0.0884 -		106	23-Dec-2021 14:02:32
88	12345	BLANK	PRO	0.0883 -		107	23-Dec-2021 14:02:50
89	12345	BLANK	PRO	0.0889 -		108	23-Dec-2021 14:03:08
90	2105079B	MULTICAL	PRO	0.4098 -		109	23-Dec-2021 14:03:26
91	2105079B	MULTICAL	PRO	0.4041 -		110	23-Dec-2021 14:03:44
92	2105079B	MULTICAL	PRO	0.4002 -		111	23-Dec-2021 14:04:02
93	12345	BLANK	CRE2	0.0021 -		127	23-Dec-2021 14:09:09
94	12345	BLANK	CRE2	0.0019 -		128	23-Dec-2021 14:09:27
95	12345	BLANK	CRE2	0.0019 -		129	23-Dec-2021 14:09:45
96	2105079B	MULTICAL	CRE2	0.0550 -		130	23-Dec-2021 14:10:03
97	2105079B	MULTICAL	CRE2	0.0549 -		131	23-Dec-2021 14:10:21
98	2105079B	MULTICAL	CRE2	0.0551 -		132	23-Dec-2021 14:10:39
99	12345	BLANK	ALPU	0.0006 -		134	23-Dec-2021 14:57:08
100	2105079B	MULTICAL	ALPU	0.0760 -		135	23-Dec-2021 14:57:25
101	12345	BLANK	UA	0.0032 -		137	23-Dec-2021 14:58:02
102	2105079B	MULTICAL	UA	0.1944 -		138	23-Dec-2021 14:58:20

Result Reprint

Report Type : Controls

08-Aug-2022

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval(3SD)
1	S062125B	ERBA NORM	GPThL	36.4 U/L		17278	08-Aug-2022 10:28:49	33.600	2.240	26.88 - 40.32
2	S062125B	ERBA NORM	GOThL	33.9 U/L		17279	08-Aug-2022 10:39:43	34.400	2.290	27.53 - 41.27
3	S062125B	ERBA NORM	ALPU	129 U/L		17280	08-Aug-2022 10:41:49	127.750	8.520	102.19 - 153.31
4	S062125B	ERBA NORM	GGT	40.9 U/L		17281	08-Aug-2022 10:43:55	41.970	2.800	33.57 - 50.37
5	S062125B	ERBA NORM	PHOS	4.32 mg/dL		17282	08-Aug-2022 10:46:02	4.730	0.240	4.01 - 5.45
6	S062125B	ERBA NORM	BIDD	0.55 mg/dL		17283	08-Aug-2022 10:48:08	0.580	0.050	0.43 - 0.73
7	S062125B	ERBA NORM	BILT	1.50 mg/dL		17284	08-Aug-2022 10:50:14	1.510	0.130	1.12 - 1.9
8	S062125B	ERBA NORM	UREA	33.6 mg/dL		17285	08-Aug-2022 10:52:21	33.970	1.700	28.87 - 39.07
9	S062125B	ERBA NORM	CRENZ	1.15 mg/dL		17286	08-Aug-2022 10:54:27	1.140	0.070	0.93 - 1.35
10	S062125B	ERBA NORM	GLU	92.3 mg/dL		17287	08-Aug-2022 10:56:34	95.720	4.780	81.38 - 110.06
11	S062125B	ERBA NORM	TRIG	116.0 mg/dL		17288	08-Aug-2022 10:58:40	113.350	5.670	96.34 - 130.36
12	S062125B	ERBA NORM	CHOL	135 mg/dL		17289	08-Aug-2022 10:58:58	137.570	6.880	116.93 - 158.21
13	S062125B	ERBA NORM	HDLC	41.7 mg/dL		17290	08-Aug-2022 10:59:34	42.860	2.860	34.28 - 51.44
14	S062125B	ERBA NORM	UA	6.0 mg/dL		17291	08-Aug-2022 11:00:52	6.080	0.300	5.18 - 6.98
15	S062125B	ERBA NORM	CA	8.2 mg/dL		17292	08-Aug-2022 11:01:10	8.090	0.400	6.89 - 9.29
16	S062125B	ERBA NORM	ALBD	3.74 mg/dL	+1SD	17561	08-Aug-2022 11:01:28	3.570	0.180	3.03 - 4.11
17	S062125B	ERBA NORM	PRO	6.38 g/dL		17562	08-Aug-2022 11:01:46	6.240	0.310	5.31 - 7.17

Result Reprint

Report Type : Calibration

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
1	12345	BLANK	GPThL	-0.0003	-		21693	8-Aug-2022 16:22:55
2	2105079B	MULTICAL	GPThL	-0.0249	-		21694	8-Aug-2022 16:23:13
3	12345	BLANK	GOTHL	-0.0008	-		21695	8-Aug-2022 16:23:31
4	2105079B	MULTICAL	GOTHL	-0.0304	-		21696	8-Aug-2022 16:23:49
5	12345	BLANK	GGT	0.0012	-		21699	8-Aug-2022 16:24:43
6	12345	BLANK	GGT	0.0014	-		21700	8-Aug-2022 16:25:01
7	12345	BLANK	GGT	0.0013	-		21701	8-Aug-2022 16:25:19
8	2105079B	MULTICAL	GGT	0.0673	-		21702	8-Aug-2022 16:25:37
9	2105079B	MULTICAL	GGT	0.0671	-		21703	8-Aug-2022 16:25:55
10	2105079B	MULTICAL	GGT	0.0674	-		21704	8-Aug-2022 16:26:13
11	12345	BLANK	PHOS	0.2745	-		21705	8-Aug-2022 16:26:31
12	12345	BLANK	PHOS	0.2713	-		21706	8-Aug-2022 16:26:49
13	12345	BLANK	PHOS	0.2717	-		21707	8-Aug-2022 16:27:08
14	2105079B	MULTICAL	PHOS	0.5539	-		21708	8-Aug-2022 16:27:26
15	2105079B	MULTICAL	PHOS	0.5524	-		21709	8-Aug-2022 16:27:44
16	2105079B	MULTICAL	PHOS	0.5534	-		21710	8-Aug-2022 16:28:02
17	12345	BLANK	BIDD	0.0001	-		21711	8-Aug-2022 16:28:20

Result Reprint

Report Type : Calibration

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
18	12345	BLANK	BIDD	0.0002 -		21712	8-Aug-2022 16:28:38
19	12345	BLANK	BIDD	0.0003 -		21713	8-Aug-2022 16:28:56
20	2105079B	MULTICAL	BIDD	0.0923 -		21714	8-Aug-2022 16:29:14
21	2105079B	MULTICAL	BIDD	0.0926 -		21715	8-Aug-2022 16:29:32
22	12345	BLANK	BIT	0.0002 -		21717	8-Aug-2022 16:30:08
23	12345	BLANK	BIT	0.0000 -		21718	8-Aug-2022 16:30:26
24	12345	BLANK	BIT	0.0001 -		21719	8-Aug-2022 16:30:44
25	12345	BLANK	UREA	-0.0053 -		21723	8-Aug-2022 16:31:56
26	12345	BLANK	UREA	-0.0052 -		21724	8-Aug-2022 16:32:14
27	12345	BLANK	CRENZ	0.0105 -		21727	8-Aug-2022 16:33:09
28	12345	BLANK	CRENZ	0.0086 -		21728	8-Aug-2022 16:33:26
29	12345	BLANK	GLU	0.0457 -		21729	8-Aug-2022 16:33:45
30	12345	BLANK	GLU	0.0464 -		21730	8-Aug-2022 16:34:03
31	12345	BLANK	GLU	0.0449 -		21731	8-Aug-2022 16:34:20
32	12345	BLANK	TRIG	0.0093 -		21732	8-Aug-2022 16:34:39
33	12345	BLANK	TRIG	0.0128 -		21733	8-Aug-2022 16:34:57
34	12345	BLANK	TRIG	0.0126 -		21734	8-Aug-2022 16:35:15

Result Reprint

Report Type : Calibration

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
35	12345	BLANK	CHOL	0.0403 -		21735	8-Aug-2022 16:35:33
36	12345	BLANK	CHOL	0.0393 -		21736	8-Aug-2022 16:35:51
37	12345	BLANK	CHOL	0.0391 -		21737	8-Aug-2022 16:36:09
38	12345	BLANK	UA	0.0043 -		21738	8-Aug-2022 16:36:27
39	12345	BLANK	CA	0.4392 -		21739	8-Aug-2022 16:36:45
40	12345	BLANK	CA	0.4374 -		21740	8-Aug-2022 16:37:03
41	12345	BLANK	CA	0.4366 -		21741	8-Aug-2022 16:37:21
42	12345	BLANK	ALBD	0.0810 -		21742	8-Aug-2022 16:37:39
43	12345	BLANK	ALBD	0.0807 -		21743	8-Aug-2022 16:37:57
44	12345	BLANK	ALBD	0.0806 -		21744	8-Aug-2022 16:38:15
45	12345	BLANK	PRO	0.0929 -		21745	8-Aug-2022 16:38:33
46	12345	BLANK	PRO	0.0898 -		21746	8-Aug-2022 16:38:51
47	12345	BLANK	PRO	0.0905 -		21747	8-Aug-2022 16:39:09