



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/02439

Page: 1 of 2

Customer Name:	ARROW DIAGNOSTICS	Customer Address:	NO: 134, GNT ROAD, REDHILLS, CHENNAI-600052.
Location:	DIAGNOSTICS	SRF No:	SBS/SRF/21-22/0079-006
Description:	MICRO PIPETTE	SRF Date:	26-04-2021
Make:	DRAGON LAB	Receipt Date:	26-04-2021
Model:	NA	Date Of calibration:	26-04-2021
Serial No:	YE174AB0078404	Date Of Due:	25-04-2022
Asset No:	NA	Issue Date:	29-04-2021
Range:	100 μ L		FIXED

Traceability Details:

S.No	Standard Used	Make	Model	Cer.No	Validity
1	Electronic Weighing Balance	A&D	GH-252	ECT 2020/0914-01	14-Dec-2022


Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MBV-03

REMARKS:

Calibration Report : Attached

Calibrated by:


K. SATHIYAMOORTHY
(Calibration Engineer)

Authorized Signatory:



Result of Calibration:


REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	μL	100	99.5	0.5%



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign:



Date: 26-04-2021