

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Calcium	VITROS 250	8.00	mg/dl	2.38
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : ARK Sens-e-lyte


Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	100.10	96.15	-1.65	-1.65
@ Sodium	mmol/L	130.40	129.30	0.32	0.32
@ Potassium	mmol/L	3.73	3.67	-0.36	-0.38

Instrument : VITROS 250

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	68.00	68.20	0.00	0.04
@ Urea	mg/dl	19.26	21.12	-1.39	-0.13
@ Creatinine	mg/dl	0.90	0.80	1.82	1.24
Bilirubin Direct	mg/dl				
@ Total Protein	g/dL	6.20	6.14	0.00	-0.81
@ Albumin	g/dL	3.60	3.70	0.00	-0.02
@ ALT	U/L	7.00	6.80	-1.57	-1.13
@ AST	U/L	22.00	22.80	0.00	-0.35
@ ALP	U/L	55.00	58.00	-0.51	-0.46
@ LDH	U/L	155.00	138.00	1.03	0.49
@ Uric Acid	mg/dl	3.80	3.86	-1.14	-1.18
! Calcium	mg/dl	8.00	7.90	2.38	0.19
@ Phosphorous Inorganic	mg/dl	3.70	4.03	-1.93	-0.05
@ Cholesterol	mg/dl	135.00	134.40	0.00	-0.40
@ Triglycerides	mg/dl	133.00	137.60	0.00	-0.21
@ HDL	mg/dl	28.00	31.40	-0.54	0.10

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	0
Evaluated Parameters	19
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070