

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
X Specific Gravity	Uriskan 100 plus	1.030	-
# Blood (urine)	Uriskan 100 plus	1+	-
# Nitrite (urine)	Uriskan 100 plus	Present	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01


Report Date : 30/12/2021

Instrument : Uriskan 100 plus

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,6,6.5
X Specific Gravity	-	1.030	1.015,1.020,1.025
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,Tr
@ Ketone (urine)	-	Negative	Negative
# Blood (urine)	-	1+	
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
# Nitrite (urine)	-	Present	

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	85.71 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report