

SHRUTI CLINICAL LAB (586)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SHRUTI CLINICAL LAB (586)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**


**Report Date : 23/05/2022**

**Instrument : Sysmex XN 350**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	4.76	4.87	-0.15	0.19
@ Total Leucocytes (WBC) count	/c.mm	2910.00	2,873.33	-0.98	-1.13
@ Haemoglobin	g/dL	15.90	16.23	-0.16	-0.02
@ MCV (Mean Corpuscular Volume)	fL	98.50	102.67	-0.51	0.19
@ MCH (Mean Corpuscular Hb)	pg	33.40	33.30	0.05	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	33.90	32.47	0.59	-0.25
@ Platelet Count	1000/ $\mu$ L	112.00	110.75	-1.49	-0.99
@ Hematocrit	%	46.90	50.07	-0.42	-0.48
@ RDW CV%	%	12.60	13.20	-0.14	0.05

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**

**PT coordinator & Technical Manager, MHL EQAS**

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report

**SHRUTI CLINICAL LAB (586)**

**Outlier And Analyte Summary Report**

**Outlier Details For Cycle No 220104 and Sample No 06**

**Report Date : 23/05/2022**

Analyte	Instrument	Result Value	Standard Unit
# Neutrophil.	-	40-70%	-
# Lymphocyte.	-	20-40%	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SHRUTI CLINICAL LAB (586)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 06**


**Report Date : 23/05/2022**

**Instrument : -**

Analyte	Standard Unit	Result Value	Accepted Value
# Neutrophil.	-	40-70%	
# Lymphocyte.	-	20-40%	
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	2
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report