

The Ultimate Calibration Services



The Quality Solutions....

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503. Phone No.: 9890715627 / 9665003365, Mail: theucs20@gmail.com, info@theultimatecalibration.com * Website: www.theultimatecalibration.com GST No.: 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CA	LIBRA	MOITA	CERTIFICATE	

Calibration Certificate No: UCS/CAL/21110172

Date of calibration: 25-11-2021

ULR No:CC302021000001588F

Name and Address of Customer:

Shruti Clinical lab

203/204, Royal Chambers, N.S Road no- 1,

Opp Club Millenium, Mumbai - 400049

Ref. No: Letter

Date: 24-11-2021

Date of Receipt: 24-11-2021 Condition on Receipt: Good

Date of Issue: 29-11-2021

Suggested Due Date: 25-11-2022

Details of Equipment:

Name: Micropipette

Make: ---

Id.No .: ---

Digital Balance

Model:---Department: ---

Least Count: 5 µl

Calibration Points: 100, 500, 1000 µl

Calibration Procedure Used: UCS/CAL/201

Details of STD equipment used:

Environmental Conditions : 23 \pm 2 °C & 50 \pm 10 % RH Name ld No

Parameter/Uncertainty UCS/CAL/058 Weigh: ±0.006 to 0.007 mg

OBSERVATIONS:

Range : 100 to 1000 µl

Certificate No MWS/CC/2021/

Sr.No.: GP30025

Accuracy: Not Specified

Location: --

Valid up to 25-02-2022

076-001

Discipline: Mechanical

Parameter: Volume

Sub Group: Mass and Volume

			Sub Group . Wass and volume	
Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
μΙ	μΙ	μΙ	μl	μl
100 to 1000	100	99.37	0.63	± 0.59
	500	499.10	0.90	± 0.55
	1000	999.10	0.90	± 0.55

Calibrated By: Holwankag

Name & Sign: Swapnil Malwankar

Authorised Signatory:

Name & Designation:

Technical Manager

1 of 2

Format No:

UCS/QF/7.8/01

Page no



The Ultimate Calibration Services

The Quality Solutions....



Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503. Phone No.: 9890715627 / 9665003365, Mail: theucs20@gmail.com, info@theultimatecalibration.com • Website: www.theultimatecalibration.com GST No.: 27AANFT7678Q1ZT 1 PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110172

Date of calibration: 25-11-2021 ULR No:CC302021000001588F

Date of Issue: 29-11-2021 Suggested Due Date: 25-11-2022

Remarks:

1. The readings noted above are average of five readings.

2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor k=2 at a confidence level of 95%.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.

The Calibration Certificate issued is for scientific and industrial purpose only.

Calibrated By:

Name & Sign: Swapnil Malwankar

Authorised Signatory: Name & Designation: V.W.Patwardhan

Technical Manager

Format No: UCS/QF/7.8/01 -----End of Certificate-----

Page no

2 of 2