

The Ultimate Calibration Services



The Quality Solutions

Date of Issue: 28-07-2022

Suggested Due Date: 25-11-2022

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.

Phone No.: 9890715627 / 9665003365, Mail: theucs20@gmail.com, info@theultimatecalibration.com • Website: www.theultimatecalibration.com

GST No.: 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CAI	JIBR A	MOIT	CERTIFIC	ATE

Calibration Certificate No: UCS/CAL/21110171A

Date of calibration: 25-11-2021

ULR No:CC302021000001587F

Name and Address of Customer: Ref. No: Letter

Shruti Clinical lab Date: 24-11-2021

203/204, Royal Chambers, N.S Road no- 1, Date of Receipt: 24-11-2021
Opp Club Millenium, Mumbai - 400049 Condition on Receipt: Good

Details of Equipment:

 Name : Micropipette
 Model : -- Sr.No.: 8017640

Id.No.: --- Department : --- Location : ---

Least Count : 0.5 μl Range : 5 to 50 μl Accuracy : Not Specified

Calibration Points: 10, 25, 50 µl

Calibration Procedure Used: UCS/CAL/201 Environmental Conditions: 23 ± 2 °C & 50 ± 10 % RH

Details of STD equipment used:

Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Balance	UCS/CAL/058	Weigh: ±0.006 mg	MWS/CC/2021/	25-02-2022
			076-001	1

OBSERVATIONS:

Discipline : Mechanical Parameter : Volume Sub Group : Mass and Volume

Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
μΙ	μΙ	μΙ	μΙ	μΙ
5 to 50	10	9.43	0.57	± 0.22
	25	24.74	0.26	± 0.22
	50	50.12	-0.12	± 0.52

Calibrated By: Plakankar Name & Sign: Swapnil Malwankar CALIBRATION SERVICO

Authorised Signatory:

Name & Designation: V.W.Patwardhan

1 of 2

Technical Manager

P.T.O.

Format No: UCS/QF/7.8/01 Page no



The Ultimate Calibration Services

The Quality Solutions....



Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.

GST No.: 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110171A

Date of calibration: 25-11-2021 ULR No:CC302021000001587F Date of Issue: 28-07-2022

Suggested Due Date: 25-11-2022

Remarks:

- 1. The readings noted above are average of five readings.
- 2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor k=2 at a confidence level of 95%.
- 3. The Certificate No. UCS/CAL/21110171A dated 28-07-2022 is an amended certificate of the Certificate No. UCS/CAL/21110171 dated 29-11-2021 for the change in Sr. No. of the instrument.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.

The Calibration Certificate issued is for scientific purpose only.

Calibrated By: Plalwankar Name & Sign: Swapnil Malwankar



Authorised Signatory: Vaturalla Name & Designation: V.W.Patwardhan

Technical Manager

Format No: UCS/QF/7.8/01

---End of Certificate-----

Page no

2 of 2