



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/22070170
Date of calibration: 28-07-2022
ULR No:CC302022000001122F
Date of Issue: 29-07-2022
Suggested Due Date: 28-07-2023
Location of calibration: Lab

Name and Address of Customer:
Shruti Clinical Laboratory
203/204, Royal Chambers, N.S Road no- 1,
Opposite Club Millenium, Juhu,
Mumbai - 400069
Ref. No: Verbal
Date: 27-07-2022
Date of Receipt: 27-07-2022
Condition on Receipt: Good

Details of Equipment :
Name : Micropipette
Make : Biosystem
ID No : ---
Least Count : 10 µl
Model : Bio one
Department : ---
Range : 100 to 1000 µl
Sr.No.: BO 3433
Location : ---
Accuracy : ± 8.0 µl
Calibration Points : 100, 500, 1000 µl
Calibration Procedure Used: UCS/CAL/201
Environmental Conditions : 23 ±2 °C & 40 to 60 % RH

Details of STD equipment used :				
Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Balance	UCS/CAL/058	Weigh: ± 0.006 to 0.007 mg	MWS/CC/2022/ 090-001	16-02-2023

The measurements are traceable to National/International standard through unbroken chain of traceability.

OBSERVATIONS :

Discipline : Mechanical		Parameter : Volume		Sub Group : Mass and Volume	
Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty	
µl	µl	µl	µl	µl	
100 to 1000	100	98.197	1.803	± 0.62	
	500	500.699	-0.699	± 0.55	
	1000	1000.512	-0.512	± 3.43	

The Expanded uncertainty is calculated at the coverage factor k=2 at a confidence level of 95.45 %

Calibrated By : *Malwankar*
Name & Sign: Swapnil Malwankar



Authorised Signatory : *Patwardhan*
Name & Designation: V.W.Patwardhan
Technical Manager



The Ultimate Calibration Services

The Quality Solutions...



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/22070170

Date of Issue: 29-07-2022

Date of calibration: 28-07-2022

Suggested Due Date: 28-07-2023

ULR No:CC302022000001122F

Location of calibration: Lab

Remarks:

1. The readings noted above are average of five readings.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.


Certificate should not be reproduced except in full without the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issuing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The Calibration Certificate issued is for scientific purpose only.



Calibrated By : *Malwankar*
Name & Sign: Swapnil Malwankar

Authorised Signatory : *Patwardhan*
Name & Designation: V.W.Patwardhan

Technical Manager

Format No: UCS/QF/7.8/01 -----End of Certificate-----

Page no 2 of 2