



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110163

Date of Issue: 28-11-2021

Date of calibration: 25-11-2021

Suggested Due Date: 25-11-2022

ULR No:CC302021000001579F

Name and Address of Customer:

Shruti Clinical lab
203/204, Royal Chambers, N.S Road no- 1,
Opp Club Millenium, Mumbai - 400049

Ref. No: Letter

Date: 24-11-2021

Date of Receipt: 24-11-2021

Condition on Receipt: Good

Details of Equipment :

Name : Digital Thermometer

Make : Aceteq

Model : ST-9283B

Sr.No.: ---

Id.No.: DT-01

Department : ---

Location : ---

Least Count : 0.1 °C

Range : -50 to 300 °C

Accuracy : Not Specified

Calibration Points : -10, -20, -25 °C (As specified by customer)

Calibration Procedure Used: UCS/CAL/301

Environmental Conditions : 25 ± 3 °C & 50 ± 15 % RH

Details of STD equipment used :

Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
4 Wire RTD Sensor	UCS/CAL/020	Temperature : ± 0.41 °C	UCS/CAL/21020001	04-02-2022
Digital Multimeter	UCS/CAL/095	Resistance: ± 0.0012 %	M-211030-16-2 Godrej & Boyce	15-11-2022

OBSERVATIONS :

Discipline : Thermal

Parameter : Temperature

Sub Group : Temperature

Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
°C	°C	°C	°C	°C
-50 to 300	-9.8	-10.41	0.61	± 0.49
	-18.9	-20.40	1.50	± 0.50
	-24.5	-25.66	1.16	± 0.50



Calibrated By : *Malwankar*
Name & Sign: Swapnil Malwankar

Authorised Signatory : *V.W. Patwardhan*
Name & Designation: V.W. Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

Page no 1 of 2 P.T.O.



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110163

Date of Issue: 28-11-2021

Date of calibration: 25-11-2021

Suggested Due Date: 25-11-2022

ULR No:CC302021000001579F

Remarks:

1. The readings noted above are average of five readings.
2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor $k=2$ at a confidence level of 95%.
3. Thermometer Length = 199 mm , Diameter = 3.3 mm , Depth Of Immersion = 30 mm.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.

Calibrated By: *Malwankar*
Name & Sign: Swapnil Malwankar



Authorised Signatory: *W. Patwardhan*
Name & Designation: V.W.Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01 -----End of Certificate-----

Page no 2 of 2