



# The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.  
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com  
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

## CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110162

Date of Issue: 28-11-2021

Date of calibration: 24-11-2021

Suggested Due Date: 24-11-2022

ULR No:CC302021000001578F

Name and Address of Customer:  
Shruti Clinical lab  
203/204, Royal Chambers, N.S Road no- 1,  
Opp Club Millenium, Mumbai - 400049

Ref. No: On Site  
Date: 24-11-2021  
Date of Receipt: 24-11-2021  
Condition on Receipt: Good

### Details of Equipment :

Name : Centrifuge

Make : Remi

Model : R- 303

Sr.No.: HGLC-9423

Id.No.: ---

Department : ---

Location : ---

Least Count : ---

Range : 0 to 3500 RPM

Accuracy : Not Specified

Calibration Points : Knob Position At 1, 2, 3, 4, 5 ( As specified by customer)

Calibration Procedure Used: UCS/CAL/207

Environmental Conditions : 25 ± 10 °C & 50 ± 15 % RH

### Details of STD equipment used :

Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Tachometer	UCS/CAL/047	Speed : ± 2.850 RPM	M-210216-2-2	17-02-2022

### OBSERVATIONS :

Discipline : Mechanical		Parameter : Speed		Sub Group : Acceleration and Speed	
Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty	
RPM	Knob Position at	RPM	RPM	RPM	
0 to 3500	1	1370	---	± 7.10	
	2	1655	---	± 7.10	
	3	2106	---	± 7.10	
	4	2580	---	± 7.10	
	5	3367	---	± 7.10	

Calibrated By : Malwankar  
Name & Sign: Swapnil Malwankar



Authorised Signatory : Patwardhan  
Name & Designation: V.W.Patwardhan  
Technical Manager

Format No: UCS/QF/7.8/01

Page no 1 of 2 P.T.O.



# The Ultimate Calibration Services

*The Quality Solutions....*



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.  
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com  
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

## CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110162

Date of Issue: 28-11-2021

Date of calibration: 24-11-2021

Suggested Due Date: 24-11-2022


ULR No:CC302021000001578F

### Remarks:

1. The readings noted above are average of five readings.
2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor  $k=2$  at a confidence level of 95%.

The readings noted are particularly for the equipment submitted for calibration.  
The reported values are valid at the time of calibration and under the controlled conditions.  
Certificate should not be reproduced except in full with the prior approval of Laboratory.  
Any handmade corrections made on the certificate, invalidates the Calibration Certificate.  
The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issuing the calibration certificate.  
All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.  
The measurements are traceable to National/International standard.

Calibrated By: Malwankar  
Name & Sign: Swapnil Malwankar



Authorised Signatory : Patwardhan  
Name & Designation: V.W.Patwardhan  
Technical Manager