




The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE				
Calibration Certificate No: UCS/CAL/21110170			Date of Issue: 29-11-2021	
Date of calibration: 25-11-2021			Suggested Due Date: 25-11-2022	
ULR No:CC302021000001586F				
Name and Address of Customer: Shruti Clinical lab 203/204, Royal Chambers, N.S Road no- 1, Opp Club Millenium, Mumbai - 400049			Ref. No: Letter Date: 24-11-2021 Date of Receipt: 24-11-2021 Condition on Receipt: Good	
Details of Equipment :				
Name : Digital Timer				
Make : ERMA	Model : DT-1045		Sr.No. : ---	
Id.No: Timer-01	Department : --		Location: ---	
Least Count : 1 sec	Range : 0 to 99 min / 60 sec		Accuracy : Not specified	
Calibration Points: 1, 15, 30 min (As specified by customer)				
Calibration Procedure Used: UCS/CAL/109			Environmental Conditions : 25 ± 3 °C & 50 ± 15 % RH	
Details of STD equipment used :				
Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Time Calibrator	UCS/CAL/059	Time: ± 0.5 sec	M-211007-22-1 Godrej & Boyce	12-10-2022
OBSERVATIONS :				
Discipline : Electro-technical		Parameter : Time	Sub Group : Time and Frequency	
Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty
min	min:sec	sec	sec	sec
0 to 99	01:00	59.543	0.457	± 0.80
	15:00	899.331	0.669	± 1.28
	30:00	1799.070	0.930	± 2.17
				
Calibrated By : <i>Malwankar</i> Name & Sign : Swapnil Malwankar			Authorised Signatory : <i>W. Patwardhan</i> Name & Designation: V.W.Patwardhan Technical Manager	

Format No: UCS/QF/7.8/01

Page no 1 of 2 P.T.O.



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/21110170

Date of Issue: 29-11-2021

Date of calibration: 25-11-2021

Suggested Due Date: 25-11-2022

ULR No:CC302021000001586F

Remarks:

1. The readings noted above are average of five readings.
2. The uncertainty expressed as Expanded uncertainty is calculated at the coverage factor $k=2$ at a confidence level of 95%.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full with the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

The measurements are traceable to National/International standard.



Calibrated By : *Swapnil Malwankar*
Name & Sign : Swapnil Malwankar

Authorised Signatory : *V.W. Patwardhan*
Name & Designation: V.W.Patwardhan
Technical Manager