



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No: CC-22-23-L-297/08

Page. No : 1 of 1

<b>Customer Name &amp; Address</b> M/s. Life Care Diagnostics 11, Vaithyanatha Iyyer, 1st Street, Madurai - 625 020.	<b>ULR No.</b>	CC272522000004792F
	<b>SRF No.</b>	VCI-22-23-L-297
	<b>Date of Receipt</b>	05.07.2022
	<b>Date of Calibration</b>	06.07.2022
	<b>Due Date for Calibration</b>	05.07.2023
	<b>Date of Issue</b>	07.07.2022

**Details of Unit Under Calibration**

<b>Description</b>	Digital Thermometer	<b>Make</b>	--
<b>Range</b>	-50 to 50 °C	<b>Model</b>	--
<b>Resolution</b>	0.1 °C	<b>Condition of UUC</b>	Good
<b>Serial Number</b>	--	<b>Instrument Location</b>	Molecular Biology Lab
<b>ID Number</b>	LCD/MDU/MOL/EQP/32	<b>Cal. At</b>	Thermal Lab

Environmental Condition		Calibration Method Used	
<b>Temperature (°C)</b>	25±4	<b>National / International Standard</b>	DKD-R-5-1
<b>Humidity (%RH)</b>	35 - 75	<b>Cal Procedure No</b>	VCI-L03-WI-THER-002

**Standard Used**

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-001& VCI-ET-DT-008	2021-22/CFC/3052/1	Aparajit Instruments	National Standards	22.03.2023

**Results of Calibration**

Sl. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Measurement Uncertainty ± (°C)
01	-30.15	-30.1	0.05	0.27
02	0.42	0.1	-0.32	0.27
03	10.56	10.1	-0.46	0.27
04	30.68	30.1	-0.58	0.27
05	50.89	50.2	-0.69	0.27

**Remarks:**

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor  $k=2$ .
- Temperature Chart : ITS-90

**Checked & Issued by:**

M. Manju  
M.Manju  
(Engineer - Calibration)

**Authorised by:**

C.Dhinakaran  
(Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com, info@venuscal.com | www.venuscal.com