



C A L I B R A T I O N C E R T I F I C A T E

This is to certify that the control calibration of **FULLY AUTO BIOCHEMISTRY ANALYZER**, Model: **QUICKEM 200**, S. No: **1250620250** . Checked at **BIOTECH DIAGNOSTIC CENTRE, NO 22C, Thiruvennainallur Road, New Water Tank Opposite, Ulunderpet 606107**. Has been successfully performed on 01/07/2022. All parameters have been successfully calibrated and validated with respective controls. Results obtained are within specified range. This calibration is valid for 12 months from the date of calibration.

Calibrated on 01/07/2022

Expired on 30/06/2023

For QUICKLAB SERVICES



A. SUBANI
Service Manager

Instrument Department

QC results

From: 09-08-2022

To: 09-08-2022

Level:

No.	Item	Lot number	Result	Check date	Check time	Flag	QC range
1	ALB	landox_445	3.8	09-08-2022	17:38:47	-1SD	2.30-5.50
2	CHOL	landox_445	181.2	09-08-2022	17:38:47	+1SD	129.00-201.00
3	CRE	landox_445	0.74	09-08-2022	17:39:59	+1SD	0.20-1.00
4	GLU	landox_445	95.00	09-08-2022	17:42:00	-1SD	80.00-120.00
5	TBIL	landox_445	0.39	09-08-2022	17:41:42	+1SD	0.10-0.50
6	TGL	landox_445	164.5	09-08-2022	17:39:59	-1SD	130.00-210.00
7	TP	landox_445	3.4	09-08-2022	17:40:31	-1SD	2.40-5.60
8	UA	landox_445	3.51	09-08-2022	17:40:48	-1SD	2.10-5.30
9	UREA	landox_445	28.7	09-08-2022	17:43:47	+1SD	13.00-37.00
