

Dated 1st April2022

Certificate Of Calibration

This is to certify that the below mentioned instrument installed at Redcliffe Diagnostics Pvt Ltd, Jaipur calibrated by us on 1st April,2022. The next calibration is due on or before 31st March,2023. The values are falling within the specified range.

Instrument Model- Ci4000
Serial No. –C403141/i1sr62612
Software Version – 9.50

For Abbott Healthcare Pvt Ltd

Prashant Mishra
Technical Service Specialist
Abbott Diagnostic Division

QC Results List Report

Operator ID:

Unreleased

System serial number: 86588

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
U425/1	1	BIORADLevel 1	BIORAD - Level 1	UIBC-L	48 ug/dL			13.08.2022 09:58
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Fe-PI	264 ug/dL			13.08.2022 09:58
U425/1	1	BIORADLevel 1	BIORAD - Level 1	UA	4.7 mg/dL			13.08.2022 09:58
U425/1	1	BIORADLevel 1	BIORAD - Level 1	ALT	28 U/L			13.08.2022 09:57
U425/1	1	BIORADLevel 1	BIORAD - Level 1	GGT	75 U/L			13.08.2022 09:57
U425/1	1	BIORADLevel 1	BIORAD - Level 1	UHDL	81 mg/dL			13.08.2022 09:57
U425/1	1	BIORADLevel 1	BIORAD - Level 1	BiID	0.44 mg/dL			13.08.2022 09:56
U425/1	1	BIORADLevel 1	BIORAD - Level 1	AST	36 U/L			13.08.2022 09:56
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Urea	35 mg/dL			13.08.2022 09:56
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Chol	260 mg/dL			13.08.2022 09:55
U425/1	1	BIORADLevel 1	BIORAD - Level 1	CreaC	2.64 mg/dL			13.08.2022 09:54
U425/1	1	BIORADLevel 1	BIORAD - Level 1	AlkP	84 U/L			13.08.2022 09:54
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Phos	3.5 mg/dL			13.08.2022 09:54
U425/1	1	BIORADLevel 1	BIORAD - Level 1	BiIT	1.0 mg/dL			13.08.2022 09:52
M990/1	1	ABBTLevel 2	ABBT - Level 2	RF	31.1 IU/mL			13.08.2022 09:52
U425/1	1	BIORADLevel 1	BIORAD - Level 1	GluC	78 mg/dL			13.08.2022 09:51
U425/1	1	BIORADLevel 1	BIORAD - Level 1	TP	6.6 g/dL			13.08.2022 09:51
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Trig	185 mg/dL			13.08.2022 09:50
U425/1	1	BIORADLevel 1	BIORAD - Level 1	CaC	9.6 mg/dL			13.08.2022 09:49
U425/1	1	BIORADLevel 1	BIORAD - Level 1	AlbG	4.0 g/dL			13.08.2022 09:48
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Na-C	147 mmol/L			13.08.2022 09:48
U425/1	1	BIORADLevel 1	BIORAD - Level 1	K-C	3.9 mmol/L			13.08.2022 09:48
U425/1	1	BIORADLevel 1	BIORAD - Level 1	Cl-C	103 mmol/L			13.08.2022 09:48

Handwritten signature
13/8/22

QC Results List Report

Unreleased

Operator ID:

System serial number: 86588

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
U425/2	1	BIORADLevel 2	BIORAD - Level 2	UIBC-L	109 ug/dL			13.08.2022 10:01
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Fe-PI	73 ug/dL			13.08.2022 10:01
U425/2	1	BIORADLevel 2	BIORAD - Level 2	UA	9.8 mg/dL			13.08.2022 10:01
U425/2	1	BIORADLevel 2	BIORAD - Level 2	ALT	104 U/L			13.08.2022 10:01
U425/2	1	BIORADLevel 2	BIORAD - Level 2	GGT	184 U/L			13.08.2022 10:00
U425/2	1	BIORADLevel 2	BIORAD - Level 2	UHDL	25 mg/dL			13.08.2022 10:00
U425/2	1	BIORADLevel 2	BIORAD - Level 2	BiID	1.90 mg/dL			13.08.2022 09:59
U425/2	1	BIORADLevel 2	BIORAD - Level 2	AST	178 U/L			13.08.2022 09:59
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Urea	102 mg/dL			13.08.2022 09:59
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Chol	106 mg/dL			13.08.2022 09:58
U425/2	1	BIORADLevel 2	BIORAD - Level 2	CreaC	6.21 mg/dL			13.08.2022 09:57
U425/2	1	BIORADLevel 2	BIORAD - Level 2	AlkP	359 U/L			13.08.2022 09:57
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Phos	7.3 mg/dL			13.08.2022 09:57
U425/2	1	BIORADLevel 2	BIORAD - Level 2	BiIT	4.5 mg/dL			13.08.2022 09:55
U425/2	1	BIORADLevel 2	BIORAD - Level 2	GluC	277 mg/dL			13.08.2022 09:54
U425/2	1	BIORADLevel 2	BIORAD - Level 2	TP	4.1 g/dL			13.08.2022 09:54
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Trig	91 mg/dL			13.08.2022 09:53
U425/2	1	BIORADLevel 2	BIORAD - Level 2	CaC	12.2 mg/dL			13.08.2022 09:52
U425/2	1	BIORADLevel 2	BIORAD - Level 2	AlbG	2.6 g/dL			13.08.2022 09:51
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Na-C	128 mmol/L			13.08.2022 09:51
U425/2	1	BIORADLevel 2	BIORAD - Level 2	K-C	6.0 mmol/L			13.08.2022 09:51
U425/2	1	BIORADLevel 2	BIORAD - Level 2	Cl-C	85 mmol/L			13.08.2022 09:51

Handwritten signature
13/8/22