



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 220804A/N-102	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 04 / 08 / 2022	<b>Ref. Std.</b>	: ISO:8655:6	
<b>CALIBRATION DUE.</b>	: 03 / 02 / 2023	<b>C.P. No.</b>	: ACS/CP/08	
<b>SRF No. &amp; DATE</b>	: 220804A/N Dt. 04.08.2022	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000191F Dt. 06.08.2022	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. First Floor in block no. B 2/9 Chitrakoot Gandhi Path Vaishali Nagar Jaipur - 302021			
<b>ITEM CALIBRATED</b>	: Micro Pipette			
	<b>Make</b>	: Thermo Scientific	<b>Range</b>	: (20 to 200) µl
	<b>L. Count</b>	:	<b>ID No.</b>	: RCL/BIO/PIP-02
	<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: SW18293
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C + 3°C } R.H. { 40% to 60% }			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023
<b>RESULTS :</b> Mechanical Discipline				
<b>Nominal Capacity</b> (µl)	<b>Graduation</b> (µl)	<b>Nominal Value</b> (µl)	<b>Volume at 20 °C</b> (µl)	<b>Error</b> (µl)
		20	20.14	(-) 0.14
200		50	50.28	(-) 0.28
		100	100.36	(-) 0.36

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 0.6 \mu\text{l}$  at coverage factor  $k=2$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

