

# MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

FF/7.8/01				Page No	1 of 1
ULR No	CC334022000014131F	Date of Calibration	05.08.2022	Date of Receipt	05.08.2022
Certificate No	MKBL/22/08/0947-001	Recom. Due Date	04.08.2023	Date of Issue	12.08.2022

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S., KUMARAN DIAGNOSTIC CENTRE, VELACHERY, CHENNAI - 42.	Description	THERMO HYGROMETER
	Make / Model	HTC - 1
	Range	-10 to 50 ° C / 10 to 99 % RH / 0.1 ° C / 1 % RH
	Serial No	---
	ID No	KDC / EQP / GEN / 05
	Manufacturer Name	---
	Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/SI. No	Certificate No	Validity
01	Temperature & Humidity Indicator with sensor	MK/CAL-43 / HTI042K152714	CRMTL/01/422101234-A5	04.07.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	25 ± 4°C	Procedure No	MKBCS - TH - 09
Humidity	30 - 75 % RH	Condition of DUC	Good

### CALIBRATION RESULTS

1.THERMAL CALIBRATION TEMPERATURE				
Parameter / Range	Standard Reading ( ° C )	DUC Reading ( ° C )	Deviation ( ° C )	Expanded Uncertainty (±) (° C)
Temperature @ 50% RH	20.16	20.1	-0.06	0.42
	30.24	30.2	-0.04	
	40.32	40.2	-0.12	
	50.40	50.3	-0.10	

Parameter / Range	Standard Reading ( %RH )	DUC Reading ( %RH )	Deviation ( %RH )	Expanded Uncertainty (±) %RH
Humidity @ 25°C	30.6	30	-0.6	1.52
	50.4	50	-0.4	
	69.8	69	-0.8	

**Remarks :**

1. The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
2. The Calibration Certificate Shall not be Reproduced Expect In Full,Without Written Approval Of The Laboratory.
3. The Recalibration Interval Should be Determined on the User Requirement.
4. The Results Stated In This Certificate Relate Only to the Item Calibrated.
5. The User Should Determine The Suitability Of The Instrument For Is Intended Use.

x-x-x-x End Of Certificate x-x-x-x

Calibrated by

Authorised by

S.Suresh Kumar  
(Calibration Engineer)



L.Magesh  
(MD/QM)