

MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

CERTIFICATE OF CALIBRATION

ULR No	CC334022000014130F	Date of Calibration	05.08.2022	Date of Receipt	05.08.2022
Certificate No	MKBL/22/08/0946-002	Recom. Due Date	04.08.2023	Date of Issue	12.08.2022

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S ., SRI VIDHYA DIAGNOSTIC CENTRE , SAIDAPET , CHENNNAI - 15 .	Description	STEM THERMOMETER
	Range / Resolution	-50 - 300°C / 0.1°C
	Make / Model	MULTI / ST - 9283 B
	Identification No	SVDC / EQP / GEN / 04
	Serial No	---
	Manufacturer Name	----
	Calibrated At	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	SSPRT	MK CAL-193	CRM TL/01/421100630-A1	18.11.2022
02	DIGITAL MULTIMETER	MK/CAL-45	EQN/CRF/2201081	23.01.2023

ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD & ACCEPTANCE LIMIT	
Temperature	25 ± 4°C	Reference Std	EURAMET/cg-08/v.02
Humidity	30 - 75 % RH	Procedure No	MKBCS - TH - 02
Condition of DUC Receipt	Good		

CALIBRATION RESULTS					
1.THERMAL CALIBRATION					
S.No	Set Value °C	Standard Reading °C	DUC Reading °C	Deviation °C	Expanded Uncertainty (±) °C
1	0	0.08	0.0	-0.08	1.75
2	60	60.12	60.2	0.08	1.75
3	120	120.15	120.3	0.15	1.75
4	180	180.18	180.4	0.22	1.75
5	240	240.26	240.5	0.24	1.75
6	300	300.30	300.5	0.20	1.75

- Remarks :**
- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
 - The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
 - The Recalibration Interval Should be Determined on the User Requirement.
 - The Results Stated In This Certificate Relate Only to the Item Calibrated.
 - The User Should Determine The Suitability Of The Instrument For Is Intended Use.

x-x-x-x- End Of Certificate -x-x-x-x

Calibrated by

S.Suresh Kumar
(Calibration Engineer)



Authorised By

L.Magesh